

LB113028


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

51915



A= 0.0000e+000

B= 1.9298e-004 slope

C= -1.4556e-002 y-intercept

Rho= 0.9999948

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%CD
0.00	0.000	0.008	0.008	115	0.000	115					1
0.05	0.050					334					-1
0.20	0.200	0.199	-0.001	1107	0.0 %	1107					-1
2.50	2.500	2.483	-0.017	12941	0.0 %	12941					0
5.00	5.000	5.018	0.018	26081	0.0 %	26081					0
7.50	7.500	7.488	-0.012	38879	0.0 %	38879					0
10.0	10.000	10.004	0.004	51915	0.0 %	51915					0

LB113088
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	115	-		0 SFAM01.1	PPB	2/10/2021 14:49	Std
	0.2 S01	1107	-		0.2 SFAM01.1	PPB	2/10/2021 14:51	Std
	2.5 S02	12941	-		2.5 SFAM01.1	PPB	2/10/2021 14:54	Std
	5 S03	26081	-		5 SFAM01.1	PPB	2/10/2021 14:56	Std
	7.5 S04	38879	-		7.5 SFAM01.1	PPB	2/10/2021 14:58	Std
	10 S05	51915	-		10 SFAM01.1	PPB	2/10/2021 15:00	Std
ICV012	ICV012	20157	3.8753 -		SFAM01.1	PPB	2/10/2021 15:03	SMPL
ICB012	ICB012	9	-0.0128 -		SFAM01.1	PPB	2/10/2021 15:05	SMPL
CCV055	CCV055	25764	4.9573 -		SFAM01.1	PPB	2/10/2021 15:08	SMPL
CCB055	CCB055	9	-0.0128 -		SFAM01.1	PPB	2/10/2021 15:10	SMPL
PB134563BL	PBS463	62	-0.0026 -		SFAM01.1	PPB	2/10/2021 15:12	SMPL
M1325-19	X3023	9285	1.7772 -		SFAM01.1	PPB	2/10/2021 15:17	SMPL
M1325-20	X3023D	8777	1.6792 -		SFAM01.1	PPB	2/10/2021 15:21	SMPL
M1325-21	X3023S	22574	4.3417 -		SFAM01.1	PPB	2/10/2021 15:26	SMPL
CCV056	CCV056	25340	4.8755 -		SFAM01.1	PPB	2/10/2021 15:28	SMPL
CCB056	CCB056	12	-0.0122 -		SFAM01.1	PPB	2/10/2021 15:30	SMPL