

LB113098


ISM02.4

INSTRUMENT ID: CV1

Linear

 μ Abs.:

44933



A= 0.0000e+000

B= 2.2315e-004 slope

C= -1.5158e-002 y-intercept

Rho= 0.9999802

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.015	0.015	134	0.000	134					8
0.05	0.050	0.054	0.004	312	0.0 %	312					1
0.20	0.200	0.199	-0.001	960	0.0 %	960					2
2.50	2.500	2.462	-0.038	11099	0.0 %	11099					1
5.00	5.000	5.035	0.035	22632	0.0 %	22632					0
7.50	7.500	7.473	-0.027	33556	0.0 %	33556					0
10.0	10.000	10.012	0.012	44933	0.0 %	44933					0

LB113098
INSTRUMENT ID :CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
0 S0		134	-		0 ISM02.4	PPB	2/11/2021 13:11	Std
0.05 S0.05		312	-	0.05	ISM02.4	PPB	2/11/2021 13:13	Std
0.2 S0.2		960	-	0.2	ISM02.4	PPB	2/11/2021 13:15	Std
2.5 S2.5		11099	-	2.5	ISM02.4	PPB	2/11/2021 13:17	Std
5 S5		22632	-	5	ISM02.4	PPB	2/11/2021 13:20	Std
7.5 S7.5		33556	-	7.5	ISM02.4	PPB	2/11/2021 13:22	Std
10 S10		44933	-	10	ISM02.4	PPB	2/11/2021 13:32	Std
ICV	ICV	17105	3.8019 -		ISM02.4	PPB	2/11/2021 13:35	SMPL
ICB	ICB	68	0 -		ISM02.4	PPB	2/11/2021 13:37	SMPL
CCV060	CCV060	22500	5.0058 -		ISM02.4	PPB	2/11/2021 13:39	SMPL
CCB060	CCB060	62	-0.0013 -		ISM02.4	PPB	2/11/2021 13:41	SMPL
PB134588BL	PBW588	52	-0.0036 -		ISM02.4	PPB	2/11/2021 13:44	SMPL
M1373-01	MEW101	-7	-0.0167 -		ISM02.4	PPB	2/11/2021 13:46	SMPL
M1373-02	MEW102	-20	-0.0196 -		ISM02.4	PPB	2/11/2021 13:48	SMPL
M1373-03	MEW103	12	-0.0125 -		ISM02.4	PPB	2/11/2021 13:50	SMPL
M1373-04	MEW103D	7	-0.0136 -		ISM02.4	PPB	2/11/2021 13:53	SMPL
M1373-05	MEW103S	4477	0.9839 -		ISM02.4	PPB	2/11/2021 13:55	SMPL
CCV061	CCV061	22395	4.9824 -		ISM02.4	PPB	2/11/2021 13:57	SMPL
CCB061	CCB061	60	-0.0018 -		ISM02.4	PPB	2/11/2021 13:59	SMPL