



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : ketankumar

Parameter: pH

Supervisor Review By : apatel

Run Number: LB113099

Slope : 98.7

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2633
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.1	4.01	02/11/2021	13:40
2	CAL2	1	Water	NA	NA	20.2	7.00	02/11/2021	13:41
3	CAL3	1	Water	NA	NA	20.1	10.02	02/11/2021	13:42
4	ICV	1	Water	NA	NA	20.2	7.00	02/11/2021	13:43
5	CCV1	1	Water	NA	NA	20.1	2.01	02/11/2021	13:44
6	M1383-01	1	Water	NA	NA	20.5	7.03	02/11/2021	13:45
7	M1383-01DUP	1	Water	NA	NA	20.6	7.05	02/11/2021	13:46
8	CCV2	1	Water	NA	NA	20.1	12.01	02/11/2021	13:47

WORKLIST(Hardcopy Internal Chain)

LB113099

WorkList Name : PH21121 WorkList ID : 146318 Department : Wet-Chemistry Date : 02-11-2021 13:34:44

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/10/2021	Water	M1383-01	pH	Cool 4 deg C	APT101	G41	WGGM-21-WW-TSF-006	02/10/2021	SM 4500-H B

Date/Time 02/11/21 13:35
 Received by: KS (Wetchem)
 Relinquished by: KS (Wetchem)

Date/Time 02/11/21 13:50
 Received by: KS (Wetchem)
 Relinquished by: KS (Wetchem)