



Analytical Summary Report

Analysis Method: 9045D

Analyst By : ketankumar

Parameter: pH

Supervisor Review By : apatel

Run Number: LB113100

Slope : 98.8

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2633
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.2	4.01	02/11/2021	14:15
2	CAL2	1	Water	NA	NA	20.3	7.00	02/11/2021	14:16
3	CAL3	1	Water	NA	NA	20.4	10.02	02/11/2021	14:17
4	ICV	1	Water	NA	NA	20.2	7.00	02/11/2021	14:18
5	CCV1	1	Water	NA	NA	20.1	2.02	02/11/2021	14:19
6	M1384-01	1	Solid	20	20	20.9	5.27	02/11/2021	14:20
7	M1384-01DUP	1	Solid	20	20	20.9	5.28	02/11/2021	14:21
8	M1384-04	1	Solid	20	20	21.0	5.42	02/11/2021	14:22
9	CCV2	1	Water	NA	NA	20.2	12.02	02/11/2021	14:23

WORKLIST(Hardcopy Internal Chain)

LB 113100

WorkList Name : PH21121-1 WorkList ID : 146320 Department : Wet-Chemistry Date : 02-11-2021 14:09:00

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/11/2021	Solid	M1384-01	pH	Cool 4 deg C	PSEG03	G61	DRUM-1	02/11/2021	9045D
02/11/2021	Solid	M1384-04	pH	Cool 4 deg C	PSEG03	G61	DRUM-2	02/11/2021	9045D

Date/Time 02/11/21 14:10
 Received by: KS (Welcher)
 Relinquished by: KS (Welcher)

Date/Time 02/11/21 14:25
 Received by: KS (Welcher)
 Relinquished by: KS (Welcher)