

LB113901


ISM02.4

INSTRUMENT ID: CV2

Linear

 μ Abs.:

4514E



A= 0.0000e+000

B= 2.1906e-004 *slope*C= 9.5980e-004 *y-intercept*

Rho= 0.9996646

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.018	0.018	77	0.000	77					1
0.05	0.050					312					1
0.20	0.200	0.203	0.003	922	0.0 %	922					2
2.50	2.500	2.456	-0.044	11205	0.0 %	11205					-2
5.00	5.000	4.940	-0.060	22546	0.0 %	22546					-1
7.50	7.500	7.692	0.192	35111	0.0 %	35111					3
10.0	10.000	9.891	-0.109	45148	0.0 %	45148					-1

LB113101
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	77	-		0 ISM02.4	PPB	2/11/2021 15:02	Std
	0.2 S0.2	922	-		0.2 ISM02.4	PPB	2/11/2021 15:05	Std
	2.5 S2.5	11205	-		2.5 ISM02.4	PPB	2/11/2021 15:07	Std
	5 S5.0	22546	-		5 ISM02.4	PPB	2/11/2021 15:09	Std
	7.5 S7.5	35111	-		7.5 ISM02.4	PPB	2/11/2021 15:12	Std
	10 S10.0	45148	-		10 ISM02.4	PPB	2/11/2021 15:14	Std
ICV	ICV	17397	3.812 -		ISM02.4	PPB	2/11/2021 15:17	SMPL
ICB	ICB	20	0.0053 -		ISM02.4	PPB	2/11/2021 15:19	SMPL
CCV062	CCV062	22560	4.943 -		ISM02.4	PPB	2/11/2021 15:21	SMPL
CCB062	CCB062	-1	0.0007 -		ISM02.4	PPB	2/11/2021 15:23	SMPL
PB134598BL	PBW598	54	0.0128 -		ISM02.4	PPB	2/11/2021 15:26	SMPL
PB134598TB	LEB598	99	0.0226 -		ISM02.4	PPB	2/11/2021 15:28	SMPL
M1378-01	MC0BJ2	37	0.0091 -		ISM02.4	PPB	2/11/2021 15:30	SMPL
M1378-02	MC0BJ3	154	0.0347 -		ISM02.4	PPB	2/11/2021 15:33	SMPL
M1378-03	MC0BJ3D	147	0.0332 -		ISM02.4	PPB	2/11/2021 15:35	SMPL
M1378-04	MC0BJ3S	4664	1.0227 -		ISM02.4	PPB	2/11/2021 15:37	SMPL
PB134599BL	PBW599	29	0.0073 -		ISM02.4	PPB	2/11/2021 15:39	SMPL
PB134599TB	LEB599	85	0.0196 -		ISM02.4	PPB	2/11/2021 15:42	SMPL
M1378-05	MC0BJ5	16	0.0045 -		ISM02.4	PPB	2/11/2021 15:44	SMPL
M1378-06	MC0BJ5D	10	0.0032 -		ISM02.4	PPB	2/11/2021 15:46	SMPL
M1378-07	MC0BJ5S	4840	1.0612 -		ISM02.4	PPB	2/11/2021 15:48	SMPL
CCV063	CCV063	21914	4.8015 -		ISM02.4	PPB	2/11/2021 15:51	SMPL
CCB063	CCB063	-7	-0.0006 -		ISM02.4	PPB	2/11/2021 15:53	SMPL