



PERCENT SOLID

Supervisor: apatel
Analyst: Alexander
Date: 2/12/2021

OVENTEMP IN Celsius(°C): 107
Time IN: 17:25
In Date: 02/11/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 105
Time OUT: 09:10
Out Date: 02/12/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB113103

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
M1384-01	DRUM-1	1	1.00	2.00	2.00	100.0	CLOTH SAMPLE
M1384-04	DRUM-2	2	1.00	2.00	2.00	100.0	CLOTH SAMPLE
M1385-01	TR-04-021121	3	1.17	9.93	8.77	86.8	
M1385-02	TR-04-021121-EPH-2	4	1.18	9.64	8.45	85.9	
M1389-01	S-1	5	1.14	9.64	7.85	78.9	
M1389-01D	S-1DUP	6	1.15	9.57	7.8	79.0	
M1389-02	S-1A	7	1.15	9.79	7.71	75.9	
M1393-01	R-LA-3-1-WC	8	1.18	9.51	7.74	78.8	
M1393-02	R-LA-3-1	9	1.18	9.71	7.77	77.3	
M1394-01	FK-BPM-03-019-ABCD	10	1.18	9.84	9.12	91.7	
M1394-02	FK-BPM-03-019-ABCD	11	1.18	9.59	8.87	91.4	
M1395-01	FK-BPM-03-020-ABCD	12	1.17	9.61	8.86	91.1	
M1395-02	FK-BPM-03-020-ABCD	13	1.17	9.84	9.05	90.9	
M1396-01	FK-BPM-04-013-ABCD	14	1.18	9.99	9.55	95.0	
M1396-02	FK-BPM-04-013-ABCD	15	1.15	9.84	9.33	94.1	
M1397-01	FK-BPM-04-014-ABCD	16	1.16	9.66	9.24	95.1	
M1397-02	FK-BPM-04-014-ABCD	17	1.15	9.63	9.12	94.0	
M1398-01	VCP-SS-1-SOIL	18	1.15	9.54	7.46	75.2	
M1398-02	VCP-SS-2-SOIL	19	1.15	9.52	7.3	73.5	
M1398-03	VCP-SS-3-SOIL	20	1.14	9.84	7.76	76.1	
M1398-04	VCP-SS-4-SOIL	21	1.15	9.58	7.21	71.9	
M1398-05	VCP-SS-5-SOIL	22	1.15	9.82	7.56	73.9	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$