


LB113109

7470A

INSTRUMENT ID: CVZ



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.002	-0.002	2	0.000	2					
0.2	0.200	0.209	0.009	812	0.0 %	812					
2.5	2.500	2.518	0.018	9695	0.0 %	9695					
5.0	5.000	4.980	-0.020	19162	0.0 %	19162					
7.5	7.500	7.461	-0.039	28708	0.0 %	28708					
10.0	10.000	10.034	0.034	38603	0.0 %	38603					

9.0D

5

1

0

0

LB113109 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 12 Feb 2021 08:51:38

Sample ID	Extended ID	μ Abs.	Conc.	Std CondMethod	Units	Date	Type	Type
0.0 - 1	50	2	-	0.00007470A	PPB	12 Feb 2021 08:54:15	S	Std
0.2 - 1	50.2	812	-	0.20007470A	PPB	12 Feb 2021 08:56:32	S	Std
2.5 - 1	52.5	9695	-	2.50007470A	PPB	12 Feb 2021 08:58:49	S	Std
5.0 - 1	55	19162	-	5.00007470A	PPB	12 Feb 2021 09:01:06	S	Std
7.5 - 1	57.5	28708	-	7.50007470A	PPB	12 Feb 2021 09:06:22	S	Std
10.0 - 1	510	38603	-	10.00007470A	PPB	12 Feb 2021 09:08:38	S	Std
ICV99 - 1	ICV99	14449	3.7542	-7470A	PPB	12 Feb 2021 09:11:28	U	SMPL
ICB99 - 1	ICB99	-82	-0.0238	-7470A	PPB	12 Feb 2021 09:13:43	U	SMPL
CCV64 - 1	CCV64	19666	5.1106	-7470A	PPB	12 Feb 2021 09:15:58	U	SMPL
CCB64 - 1	CCB64	-72	-0.0212	-7470A	PPB	12 Feb 2021 09:18:13	U	SMPL
CRA - 1	CRA	788	0.2024	-7470A	PPB	12 Feb 2021 09:20:27	U	SMPL
HighStd - 1	HighStd	39026	10.1441	-7470A	PPB	12 Feb 2021 09:22:43	U	SMPL
ChkStd - 1	ChkStd	27287	7.0920	-7470A	PPB	12 Feb 2021 09:24:58	U	SMPL
PB134589BL - 1	PBW	-106	-0.0300	-7470A	PPB	12 Feb 2021 09:27:14	U	SMPL
PB134589BS - 1	LCSW	15575	4.0470	-7470A	PPB	12 Feb 2021 09:29:30	U	SMPL
M1348-03 - 1	AS-CON	-45	-0.0141	-7470A	PPB	12 Feb 2021 09:31:46	U	SMPL
M1348-03DUP - 1	AS-CONDUP	6	-0.0009	-7470A	PPB	12 Feb 2021 09:34:03	U	SMPL
M1348-03MS - 1	AS-CONMS	15521	4.0329	-7470A	PPB	12 Feb 2021 09:36:20	U	SMPL
M1348-03MSD - 1	AS-CONMSD	15652	4.0670	-7470A	PPB	12 Feb 2021 09:38:37	U	SMPL
M1367-01 - 1	EXTERIOR-PAINT-CHIPS	121	0.0290	-7470A	PPB	12 Feb 2021 09:40:53	U	SMPL
CCV65 - 1	CCV65	19382	5.0367	-7470A	PPB	12 Feb 2021 09:43:08	U	SMPL
CCB65 - 1	CCB65	-65	-0.0193	-7470A	PPB	12 Feb 2021 09:45:24	U	SMPL
M1368-01 - 1	EXTERIOR-PAINT-CHIPS	206	0.0511	-7470A	PPB	12 Feb 2021 09:47:39	U	SMPL
PB134589TB - 1	PB134574TB	16	0.0017	-7470A	PPB	12 Feb 2021 09:49:55	U	SMPL
M1348-03LX5 - 1		8	-0.0004	-7470A	PPB	12 Feb 2021 09:52:11	U	SMPL
M1348-03A - 1		15220	3.9547	-7470A	PPB	12 Feb 2021 09:54:27	U	SMPL
CCV66 - 1	CCV66	19382	5.0367	-7470A	PPB	12 Feb 2021 09:56:43	U	SMPL
CCB66 - 1	CCB66	-54	-0.0165	-7470A	PPB	12 Feb 2021 09:58:59	U	SMPL