


LB113110

ISM02.4



Linear

A= 0.0000e+000

B= 2.0132e-004 slope

C= -2.5309e-002 y-intercept

Rho= 0.9999580

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.009	-0.009	80	0.000	80					=
0.05	0.050					312					-7
0.20	0.200	0.186	-0.014	1049	0.0 %	1049					0
2.50	2.500	2.495	-0.005	12519	0.0 %	12519					1
5.00	5.000	5.028	0.028	25100	0.0 %	25100					1
7.50	7.500	7.554	0.054	37647	0.0 %	37647					1
10.0	10.000	9.946	-0.054	49531	0.0 %	49531					1

LB113110

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	80	-		0 ISM02.4	PPB	2/12/2021 10:12	Std
	0.2 S0.2	1049	-		0.2 ISM02.4	PPB	2/12/2021 10:14	Std
	2.5 S2.5	12519	-		2.5 ISM02.4	PPB	2/12/2021 10:17	Std
	5 S5.0	25100	-		5 ISM02.4	PPB	2/12/2021 10:19	Std
	7.5 S7.5	37647	-		7.5 ISM02.4	PPB	2/12/2021 10:21	Std
	10 S10.0	49531	-		10 ISM02.4	PPB	2/12/2021 10:24	Std
ICV	ICV	19393	3.879 -		ISM02.4	PPB	2/12/2021 10:26	SMPL
ICB	ICB	-14	-0.0281 -		ISM02.4	PPB	2/12/2021 10:29	SMPL
CCV067	CCV067	24975	5.0027 -		ISM02.4	PPB	2/12/2021 10:31	SMPL
CCB067	CCB067	-12	-0.0277 -		ISM02.4	PPB	2/12/2021 10:33	SMPL
PB134604BL	PBS604	50	-0.0152 -		ISM02.4	PPB	2/12/2021 10:35	SMPL
M1309-01	MCOA32	2817	0.5418 -		ISM02.4	PPB	2/12/2021 10:38	SMPL
M1309-02	MCOA33	778	0.1313 -		ISM02.4	PPB	2/12/2021 10:40	SMPL
M1309-03	MCOA34	838	0.1434 -		ISM02.4	PPB	2/12/2021 10:42	SMPL
M1309-04	MCOA56	790	0.1337 -		ISM02.4	PPB	2/12/2021 10:44	SMPL
M1309-05	MCOA57	1861	0.3494 -		ISM02.4	PPB	2/12/2021 10:47	SMPL
M1309-06	MCOA39	3867	0.7532 -		ISM02.4	PPB	2/12/2021 10:49	SMPL
M1309-07	MCOA58	168	0.0085 -		ISM02.4	PPB	2/12/2021 10:51	SMPL
M1309-08	MCOA58D	232	0.0214 -		ISM02.4	PPB	2/12/2021 10:53	SMPL
M1309-09	MCOA58S	12536	2.4985 -		ISM02.4	PPB	2/12/2021 10:56	SMPL
CCV068	CCV068	24706	4.9486 -		ISM02.4	PPB	2/12/2021 10:58	SMPL
CCB068	CCB068	-15	-0.0283 -		ISM02.4	PPB	2/12/2021 11:00	SMPL