



Analytical Summary Report

Analysis Method: 9040C

Analyst By : Alexander

Parameter: pH

Supervisor Review By : apatel

Run Number: LB113118

Slope : 98.4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2633
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	02/12/2021	13:13
2	CAL2	1	Water	NA	NA	20.2	7.01	02/12/2021	13:15
3	CAL3	1	Water	NA	NA	20.2	10.03	02/12/2021	13:16
4	ICV	1	Water	NA	NA	20.2	7.00	02/12/2021	13:18
5	CCV1	1	Water	NA	NA	20.4	2.01	02/12/2021	13:20
6	M1404-02	1	Water	NA	NA	19.7	6.20	02/12/2021	13:23
7	M1404-02DUP	1	Water	NA	NA	19.7	6.19	02/12/2021	13:25
8	M1404-03	1	Water	NA	NA	19.6	6.29	02/12/2021	13:27
9	M1404-04	1	Water	NA	NA	19.6	6.32	02/12/2021	13:30
10	M1404-05	1	Water	NA	NA	19.5	5.85	02/12/2021	13:32
11	CCV2	1	Water	NA	NA	20.4	12.02	02/12/2021	13:35

18113118 WORKLIST(Hardcopy Internal Chain)

WorkList Name : pH-21221 WorkList ID : 146357 Department : Wet-Chemistry Date : 02-12-2021 13:08:49

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/12/2021	Water	M1404-02	pH	Cool 4 deg C	PSEG03	H11	NWB-1579	02/12/2021	9040C
02/12/2021	Water	M1404-03	pH	Cool 4 deg C	PSEG03	H11	NWB-1580	02/12/2021	9040C
02/12/2021	Water	M1404-04	pH	Cool 4 deg C	PSEG03	H11	NWB-1581	02/12/2021	9040C
02/12/2021	Water	M1404-05	pH	Cool 4 deg C	PSEG03	H11	NWB-1582	02/12/2021	9040C

Date/Time 2/12/2021 13:10
 Received by: AM (wet-chem)
 Relinquished by: KS (wet-chem)

Date/Time 2/12/2021 13:45
 Received by: KS (wet-chem)
 Relinquished by: AM (wet-chem)