


ISM02.4

LB113119
INSTRUMENT ID: CV2 μ Abs.:

4068E



Linear

A= 0.0000e+000

B= 2.4617e-004 *slope*C= -1.2199e-002 *y-intercept*

Rho= 0.9999168

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.028	0.028	162	0.000	162					
0.05	0.050					312					
0.20	0.200	0.210	0.010	901	0.0 %	901					
2.50	2.500	2.497	-0.003	10195	0.0 %	10195					
5.00	5.000	4.903	-0.097	19968	0.0 %	19968					
7.50	7.500	7.558	0.058	30752	0.0 %	30752					
10.0	10.000	10.004	0.004	40688	0.0 %	40688					

0/0D
= 50-210

LB113119
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	162	-		0 ISM02.4	PPB	2/12/2021 14:22	Std
	0.2 S0.2	901	-		0.2 ISM02.4	PPB	2/12/2021 14:27	Std
	2.5 S2.5	10195	-		2.5 ISM02.4	PPB	2/12/2021 14:29	Std
	5 S5.0	19968	-		5 ISM02.4	PPB	2/12/2021 14:31	Std
	7.5 S7.5	30752	-		7.5 ISM02.4	PPB	2/12/2021 14:34	Std
	10 S10.0	40688	-		10 ISM02.4	PPB	2/12/2021 14:39	Std
ICV	ICV	15732	3.8605 -		ISM02.4	PPB	2/12/2021 14:41	SMPL
ICB	ICB	42	-0.0019 -		ISM02.4	PPB	2/12/2021 14:44	SMPL
CCV072	CCV072	20056	4.925 -		ISM02.4	PPB	2/12/2021 14:46	SMPL
CCB072	CCB072	45	-0.0011 -		ISM02.4	PPB	2/12/2021 14:48	SMPL
PB134622BL	PBW622	22	-0.0068 -		ISM02.4	PPB	2/12/2021 14:50	SMPL
M1312-01	MC0A43	-11	-0.0149 -		ISM02.4	PPB	2/12/2021 14:53	SMPL
M1312-02	MC0A54	-21	-0.0174 -		ISM02.4	PPB	2/12/2021 14:55	SMPL
M1312-03	MC0AA2	-5	-0.0134 -		ISM02.4	PPB	2/12/2021 14:57	SMPL
M1312-04	MC0AA3	-4	-0.0132 -		ISM02.4	PPB	2/12/2021 14:59	SMPL
M1312-05	MC0AA4	-11	-0.0149 -		ISM02.4	PPB	2/12/2021 15:02	SMPL
M1312-06	MC0AA1	-10	-0.0147 -		ISM02.4	PPB	2/12/2021 15:04	SMPL
M1312-07	MC0A19	-9	-0.0144 -		ISM02.4	PPB	2/12/2021 15:06	SMPL
M1312-08	MC0A45	5	-0.011 -		ISM02.4	PPB	2/12/2021 15:08	SMPL
M1312-09	MC0A46	3	-0.0115 -		ISM02.4	PPB	2/12/2021 15:11	SMPL
M1312-10	MC0A47	-20	-0.0171 -		ISM02.4	PPB	2/12/2021 15:13	SMPL
M1312-11	MC0A48	-23	-0.0179 -		ISM02.4	PPB	2/12/2021 15:15	SMPL
M1312-12	MC0A49	-16	-0.0161 -		ISM02.4	PPB	2/12/2021 15:18	SMPL
M1312-13	MC0A50	-9	-0.0144 -		ISM02.4	PPB	2/12/2021 15:20	SMPL
M1312-14	MC0A16	-27	-0.0188 -		ISM02.4	PPB	2/12/2021 15:22	SMPL
M1312-15	MC0AA5	-12	-0.0152 -		ISM02.4	PPB	2/12/2021 15:24	SMPL
M1312-16	MC0AA6	-11	-0.0149 -		ISM02.4	PPB	2/12/2021 15:27	SMPL
M1312-17	MC0AA6D	-22	-0.0176 -		ISM02.4	PPB	2/12/2021 15:32	SMPL
M1312-18	MC0AA6S	4061	0.9875 -		ISM02.4	PPB	2/12/2021 15:34	SMPL
M1312-19	MC0AA7	63	0.0033 -		ISM02.4	PPB	2/12/2021 15:36	SMPL
CCV073	CCV073	18614	4.57 -		ISM02.4	PPB	2/12/2021 15:38	SMPL
CCB073	CCB073	30	-0.0048 -		ISM02.4	PPB	2/12/2021 15:41	SMPL
M1312-20	MC0AA8	-24	-0.0181 -		ISM02.4	PPB	2/12/2021 15:43	SMPL
M1312-21	MC0A59	-7	-0.0139 -		ISM02.4	PPB	2/12/2021 15:45	SMPL
M1312-22	MC0A60	-15	-0.0159 -		ISM02.4	PPB	2/12/2021 15:47	SMPL
CCV074	CCV074	19125	4.6958 -		ISM02.4	PPB	2/12/2021 15:50	SMPL
CCB074	CCB074	40	-0.0024 -		ISM02.4	PPB	2/12/2021 15:52	SMPL