

LB113123


7471B

INSTRUMENT ID: CV2

Linear

μ Abs.:

49022



A= 0.0000e+000

B= 2.0230e-004

C= -2.6069e-002

Rho= 0.9997540

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.018	-0.018	41	0.000	41					-4
0.2	0.200	0.192	-0.008	1078	0.0 %	1078					-3
2.5	2.500	2.427	-0.073	12128	0.0 %	12128					3
5.0	5.000	5.129	0.129	25482	0.0 %	25482					1
7.5	7.500	7.578	0.078	37587	0.0 %	37587					-1
10.0	10.000	9.891	-0.109	49022	0.0 %	49022					

LB113123 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 12 Feb 2021 16:10:56					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	41	-	0.0000	7471B	PPB	12 Feb 2021 16:12:25	S	Std
0.2 - 1	50.2	1078	-	0.2000	7471B	PPB	12 Feb 2021 16:14:42	S	Std
2.5 - 1	52.5	12128	-	2.5000	7471B	PPB	12 Feb 2021 16:16:59	S	Std
5.0 - 1	55	25482	-	5.0000	7471B	PPB	12 Feb 2021 16:19:16	S	Std
7.5 - 1	57.5	37587	-	7.5000	7471B	PPB	12 Feb 2021 16:21:33	S	Std
10.0 - 1	510	49022	-	10.0000	7471B	PPB	12 Feb 2021 16:28:51	S	Std
ICV02 - 1	ICV02	19402	3.8990	-7471B		PPB	12 Feb 2021 16:31:33	U	SMPL
ICB02 - 1	ICB02	-47	-0.0356	-7471B		PPB	12 Feb 2021 16:33:48	U	SMPL
CCV75 - 1	CCV75	25228	5.0777	-7471B		PPB	12 Feb 2021 16:36:03	U	SMPL
CCB75 - 1	CCB75	-68	-0.0398	-7471B		PPB	12 Feb 2021 16:38:18	U	SMPL
CRA - 1	CRA	1033	0.1829	-7471B		PPB	12 Feb 2021 16:40:34	U	SMPL
HighStd - 1	HighStd	48524	9.7905	-7471B		PPB	12 Feb 2021 16:42:49	U	SMPL
ChkStd - 1	ChkStd	34423	6.9379	-7471B		PPB	12 Feb 2021 16:45:05	U	SMPL
PB134621BL - 1	PBS	-124	-0.0512	-7471B		PPB	12 Feb 2021 16:47:22	U	SMPL
PB134621BS - 1	LCSS	21561	4.3358	-7471B		PPB	12 Feb 2021 16:49:39	U	SMPL
M1385-01 - 1	TR-04-021121	820	0.1398	-7471B		PPB	12 Feb 2021 16:51:56	U	SMPL
M1389-01 - 1	S-1	1008	0.1779	-7471B		PPB	12 Feb 2021 16:54:13	U	SMPL
M1393-01 - 1	R-LA-3-1-WC	1395	0.2561	-7471B		PPB	12 Feb 2021 16:56:30	U	SMPL
M1394-01 - 1	FK-BPM-03-019-ABCD	1578	0.2932	-7471B		PPB	12 Feb 2021 16:58:47	U	SMPL
M1395-01 - 1	FK-BPM-03-020-ABCD	1741	0.3261	-7471B		PPB	12 Feb 2021 17:01:02	U	SMPL
CCV76 - 1	CCV76	24586	4.9478	-7471B		PPB	12 Feb 2021 17:03:17	U	SMPL
CCB76 - 1	CCB76	-53	-0.0368	-7471B		PPB	12 Feb 2021 17:05:33	U	SMPL
M1396-01 - 1	FK-BPM-04-013-ABCD	802	0.1362	-7471B		PPB	12 Feb 2021 17:07:48	U	SMPL
M1397-01 - 1	FK-BPM-04-014-ABCD	812	0.1382	-7471B		PPB	12 Feb 2021 17:10:03	U	SMPL
M1397-01DUP - 1	FK-BPM-04-014-ABCDDUP	953	0.1667	-7471B		PPB	12 Feb 2021 17:12:19	U	SMPL
M1397-01MS - 1	FK-BPM-04-014-ABCDMS	19159	3.8499	-7471B		PPB	12 Feb 2021 17:14:35	U	SMPL
M1397-01MSD - 1	FK-BPM-04-014-ABCDMSD	18917	3.8009	-7471B		PPB	12 Feb 2021 17:16:51	U	SMPL
M1397-01LX5 - 1		104	-0.0050	-7471B		PPB	12 Feb 2021 17:19:07	U	SMPL
M1397-01A - 1		20053	4.0307	-7471B		PPB	12 Feb 2021 17:21:23	U	SMPL
CCV77 - 1	CCV77	24352	4.9004	-7471B		PPB	12 Feb 2021 17:23:41	U	SMPL
CCB77 - 1	CCB77	-63	-0.0388	-7471B		PPB	12 Feb 2021 17:25:58	U	SMPL