



Analytical Summary Report

Analysis Method: 9040C

Analyst By : Alexander

Parameter: pH

Supervisor Review By : apatel

Run Number: LB113124

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2633
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	02/12/2021	15:57
2	CAL2	1	Water	NA	NA	20.3	7.01	02/12/2021	15:59
3	CAL3	1	Water	NA	NA	20.3	10.03	02/12/2021	16:00
4	ICV	1	Water	NA	NA	20.3	6.99	02/12/2021	16:02
5	CCV1	1	Water	NA	NA	20.2	2.01	02/12/2021	16:05
6	M1401-01	1	Water	NA	NA	19.7	5.41	02/12/2021	16:07
7	M1401-01DUP	1	Water	NA	NA	19.7	5.39	02/12/2021	16:10
8	CCV2	1	Water	NA	NA	20.2	12.02	02/12/2021	16:13

LB113124
 WORKLIST(Hardcopy Internal Chain)

WorkList Name : pH-21221-2 WorkList ID : 146358 Department : Wet-Chemistry Date : 02-12-2021 13:08:52

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/12/2021	Water	M1401-01	pH	Cool 4 deg C	PSEG03	H11	41702	02/12/2021	9040C

Date/Time 2/12/2021 15:50
 Received by: AM (wet-chem)
 Relinquished by: KS (wet-chem)

Date/Time 2/12/2021 16:50
 Received by: KS (wet-chem)
 Relinquished by: AM (wet-chem)