



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Alexander

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB113125

Slope : 99.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2633
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	02/12/2021	15:57
2	CAL2	1	Water	NA	NA	20.3	7.01	02/12/2021	15:59
3	CAL3	1	Water	NA	NA	20.3	10.03	02/12/2021	16:00
4	ICV	1	Water	NA	NA	20.3	6.99	02/12/2021	16:02
5	CCV1	1	Water	NA	NA	20.2	2.01	02/12/2021	16:05
6	M1400-01	1	Solid	20.02	20	19.9	8.55	02/12/2021	16:20
7	M1400-01DUP	1	Solid	20.02	20	19.9	8.59	02/12/2021	16:22
8	CCV2	1	Water	NA	NA	20.02	12.02	02/12/2021	16:25

LB113125
WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity-21221 WorkList ID : 146359 Department : Wet-Chemistry Date : 02-12-2021 15:02:15

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/12/2021	Solid	M1400-01	Corrosivity	Cool 4 deg C	PSEG03	H12	BIN-2	02/12/2021	9045D

Date/Time 2/12/2021 15:50
Received by: AM (wet-chem)
Relinquished by: KS (wet-chem)

Date/Time 2/12/2021 16:50
Received by: KS (wet-chem)
Relinquished by: AM (wet-chem)