



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB113126

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: FLASHPONT

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.00	80.6	80.6	02/15/2021	09:35
2	M1401-01		1	10	100.00	>212.0	>212.0	02/15/2021	10:05
3	M1401-01DUP		1	10	100.00	>212.0	>212.0	02/15/2021	10:40
4	M1404-01		1	9	100.00	>212.0	>212.0	02/15/2021	11:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp21221 WorkList ID : 146341 Department : Wet-Chemistry Date : 02-12-2021 08:53:35

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/19/2021	Water	M1401-01	Flash Point	Cool 4 deg C	PSEG03	H11	41702	02/12/2021	1010A
02/26/2021	Water	M1404-01	Flash Point	Cool 4 deg C	PSEG03	H11	NWB-1579-1582	02/12/2021	1010A

Date/Time 02/15/21 09:05
Received by: KS (Welchen)
Relinquished by: RC CEXT-Lab

Date/Time 02/15/21 11:15
Received by: RC CEXT-Lab
Relinquished by: KS (Welchen)