


LB113128

7470A

INSTRUMENT ID: CV2

 μ Abs.:

35990



A= 0.0000e+000

B= 2.7949e-004

C= -1.7150e-002

Rho= 0.9998474

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.002	0.002	70	0.000	70					-
0.2	0.200	0.171	-0.029	673	0.0 %	673					-15
2.5	2.500	2.523	0.023	9088	0.0 %	9088					1
5.0	5.000	5.084	0.084	18250	0.0 %	18250					2
7.5	7.500	7.378	-0.122	26460	0.0 %	26460					-2
10.0	10.000	10.042	0.042	35990	0.0 %	35990					0

LB113128 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 15 Feb 2021 09:46:32

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	70	-	0.0000	7470A	PPB	15 Feb 2021 10:21:13	S	Std
0.2 - 1	50.2	673	-	0.2000	7470A	PPB	15 Feb 2021 10:23:31	S	Std
2.5 - 1	52.5	9088	-	2.5000	7470A	PPB	15 Feb 2021 10:25:48	S	Std
5.0 - 1	55	18250	-	5.0000	7470A	PPB	15 Feb 2021 10:28:04	S	Std
7.5 - 1	57.5	26460	-	7.5000	7470A	PPB	15 Feb 2021 10:30:21	S	Std
10.0 - 1	510	35990	-	10.0000	7470A	PPB	15 Feb 2021 10:37:44	S	Std
ICV03 - 1	ICV03	15517	4.3198	-	7470A	PPB	15 Feb 2021 10:45:31	U	SMPL
ICB03 - 1	ICB03	-74	-0.0378	-	7470A	PPB	15 Feb 2021 10:47:46	U	SMPL
CCV78 - 1	CCV78	17823	4.9643	-	7470A	PPB	15 Feb 2021 10:50:01	U	SMPL
CCB78 - 1	CCB78	-77	-0.0387	-	7470A	PPB	15 Feb 2021 10:52:16	U	SMPL
CRA - 1	CRA	652	0.1651	-	7470A	PPB	15 Feb 2021 10:54:31	U	SMPL
HighStd - 1	HighStd	34520	9.6310	-	7470A	PPB	15 Feb 2021 10:56:46	U	SMPL
ChkStd - 1	ChkStd	23804	6.6359	-	7470A	PPB	15 Feb 2021 10:59:01	U	SMPL
PB134627BL - 1	PBW	-113	-0.0487	-	7470A	PPB	15 Feb 2021 11:01:18	U	SMPL
PB134627BS - 1	LCSW	13663	3.8016	-	7470A	PPB	15 Feb 2021 11:03:34	U	SMPL
M1384-02 - 1	DRUM-1	-68	-0.0362	-	7470A	PPB	15 Feb 2021 11:05:50	U	SMPL
M1384-02DUP - 1	DRUM-1DUP	-10	-0.0199	-	7470A	PPB	15 Feb 2021 11:08:07	U	SMPL
M1384-02MS - 1	DRUM-1MS	12762	3.5498	-	7470A	PPB	15 Feb 2021 11:10:24	U	SMPL
M1384-02MSD - 1	DRUM-1MSD	14305	3.9810	-	7470A	PPB	15 Feb 2021 11:12:41	U	SMPL
M1384-05 - 1	DRUM-2	-55	-0.0325	-	7470A	PPB	15 Feb 2021 11:14:56	U	SMPL
CCV79 - 1	CCV79	17380	4.8405	-	7470A	PPB	15 Feb 2021 11:17:12	U	SMPL
CCB79 - 1	CCB79	-71	-0.0370	-	7470A	PPB	15 Feb 2021 11:19:27	U	SMPL
PB134627TB - 1	PB134605TB	-17	-0.0219	-	7470A	PPB	15 Feb 2021 11:21:42	U	SMPL
M1384-02LX5 - 1		-19	-0.0225	-	7470A	PPB	15 Feb 2021 11:23:57	U	SMPL
M1384-02A - 1		13848	3.8533	-	7470A	PPB	15 Feb 2021 11:26:13	U	SMPL
CCV80 - 1	CCV 80	16550	4.6085	-	7470A	PPB	15 Feb 2021 11:28:29	U	SMPL
CCB80 - 1	CCB 80	-89	-0.0420	-	7470A	PPB	15 Feb 2021 11:30:45	U	SMPL