



PERCENT SOLID

Supervisor: apatel
Analyst: Alexander
Date: 2/16/2021

OVENTEMP IN Celsius(°C): 107
Time IN: 12:25
In Date: 02/15/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 105
Time OUT: 09:15
Out Date: 02/16/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB113131

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
M1371-01	DBH02	1	1.18	9.83	8.65	86.4	
M1371-02	DBH14	2	1.16	9.79	8.45	84.5	
M1371-03	DBH15	3	1.18	9.51	7.96	81.4	
M1371-04	DBH16	4	1.18	9.52	8.11	83.1	
M1371-05	DBH16MS	5	1.18	9.52	8.11	83.1	
M1371-06	DBH16MSD	6	1.18	9.52	8.11	83.1	
M1371-07	DBH17	7	1.17	9.87	8.39	83.0	
M1371-08	DBH18	8	1.13	9.66	8.29	83.9	
M1371-09	DBH19	9	1.13	9.57	8.01	81.5	
M1371-10	DBH20	10	1.17	9.79	8.49	84.9	
M1371-11	DBH21	11	1.15	9.62	8.48	86.5	
M1371-12	DBH22	12	1.17	9.74	8.38	84.1	
M1371-13	DBH24	13	1.15	9.79	7.72	76.0	
M1371-14	DBH25	14	1.17	9.6	8.53	87.3	
M1371-15	DBH26	15	1.18	9.98	9.26	91.8	
M1371-16	DBH30	16	1.17	9.86	9.84	99.8	
M1371-17	DBH31	17	1.17	9.83	9.82	99.9	
M1371-18	DBH32	18	1.17	9.67	9.65	99.8	
M1371-19	DBH33	19	1.17	9.88	9.85	99.7	
M1371-20	DBH35	20	1.16	9.64	8.55	87.1	
M1371-21	VHBLK01	21	1.00	2.00	2.00	100.0	
M1371-23	DBGX8	22	1.16	9.81	8.65	86.6	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$