

L3113133


ISM02.4

INSTRUMENT ID: CX1

Linear

μ Abs.:

40757



A= 0.0000e+000

B= 2.4499e-004 slope

C= -4.6482e-002 $\neq$ intercept

Rho= 0.9998886

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.003	0.003	203	0.000	203					1
0.05	0.050					312					1
0.20	0.200	0.201	0.001	1011	0.0 %	1011					-3
2.50	2.500	2.426	-0.074	10093	0.0 %	10093					2
5.00	5.000	5.080	0.080	20926	0.0 %	20926					1
7.50	7.500	7.550	0.050	31009	0.0 %	31009					1
10.0	10.000	9.939	-0.061	40757	0.0 %	40757					-1

LB113133
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	203	-		0 ISM02.4	PPB	2/15/2021 12:23	Std
	0.2 S0.2	1011	-		0.2 ISM02.4	PPB	2/15/2021 12:25	Std
	2.5 S2.5	10093	-		2.5 ISM02.4	PPB	2/15/2021 12:27	Std
	5 S5.0	20926	-		5 ISM02.4	PPB	2/15/2021 12:30	Std
	7.5 S7.5	31009	-		7.5 ISM02.4	PPB	2/15/2021 12:32	Std
	10 S10.0	40757	-		10 ISM02.4	PPB	2/15/2021 12:34	Std
ICV	ICV	15265	3.6933 -		ISM02.4	PPB	2/15/2021 12:37	SMPL
ICB	ICB	128	-0.0151 -		ISM02.4	PPB	2/15/2021 12:40	SMPL
CCV081	CCV081	20519	4.9805 -		ISM02.4	PPB	2/15/2021 12:42	SMPL
CCB081	CCB081	115	-0.0183 -		ISM02.4	PPB	2/15/2021 12:44	SMPL
PB134641BL	PBW641	-8	-0.0484 -		ISM02.4	PPB	2/15/2021 12:46	SMPL
M1416-01	MBG0T8	-27	-0.0531 -		ISM02.4	PPB	2/15/2021 12:49	SMPL
CCV082	CCV082	20123	4.8835 -		ISM02.4	PPB	2/15/2021 12:51	SMPL
CCB082	CCB082	119	-0.0173 -		ISM02.4	PPB	2/15/2021 12:53	SMPL