



Analytical Summary Report

Analysis Method: 9040C

Analyst By : Alexander

Parameter: pH

Supervisor Review By : apatel

Run Number: LB113134

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
MANGANOUS SULFATE SOLUTION-364	W2767
PH 10.01 BUFFER,COLOR CD 475ML	W2633
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	02/15/2021	12:20
2	CAL2	1	Water	NA	NA	20.1	7.01	02/15/2021	12:21
3	CAL3	1	Water	NA	NA	20.1	10.03	02/15/2021	12:23
4	ICV	1	Water	NA	NA	20.1	6.98	02/15/2021	12:25
5	CCV1	1	Water	NA	NA	20.3	2.01	02/15/2021	12:28
6	M1419-01	1	Water	NA	NA	18.8	4.74	02/15/2021	12:31
7	M1419-01DUP	1	Water	NA	NA	18.8	4.72	02/15/2021	12:33
8	M1419-02	1	Water	NA	NA	18.6	4.72	02/15/2021	12:35
9	M1419-03	1	Water	NA	NA	18.6	4.81	02/15/2021	12:38
10	M1421-01	1	Water	NA	NA	18.7	4.97	02/15/2021	12:40
11	CCV2	1	Water	NA	NA	20.2	12.02	02/15/2021	12:42

WORKLIST(Hardcopy Internal Chain)

LB113114

WorkList Name : pf21221 WorkList ID : 146340 Department : Wet-Chemistry Date : 02-12-2021 08:52:51

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/11/2021	Solid	M1385-01	Paint Filter	Cool 4 deg C	PSEG05	G61	TR-04-021121	02/11/2021	9095B
02/11/2021	Solid	M1393-01	Paint Filter	Cool 4 deg C	PSEG03	G51	R-LA-3-1-WC	02/11/2021	9095B

Date/Time 02/12/21 08:55
 Received by: KS (Wekchow)
 Relinquished by: RC (Ext. Lab)

Date/Time 02/12/21 09:35
 Received by: RC (Ext. Lab)
 Relinquished by: KS (Wekchow)