


LB113137

ISM02.4

INSTRUMENT ID: CV2

 μ Abs.:

43513



Linear

A= 0.0000e+000

B= 2.2916e-004 *slope*C= -1.9148e-002 *y-intercept*

Rho= 0.9998912

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.010	-0.010	40	0.000	40					
0.05	0.050					312					
0.20	0.200	0.166	-0.034	810	0.0 %	810					
2.50	2.500	2.485	-0.015	10926	0.0 %	10926					
5.00	5.000	5.118	0.118	22417	0.0 %	22417					
7.50	7.500	7.488	-0.012	32761	0.0 %	32761					
10.0	10.000	9.952	-0.048	43513	0.0 %	43513					

11
 -17
 -1
 2
 0
 0

LB113137

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	40	-		0 ISM02.4	PPB	2/15/2021 14:00	Std
	0.2 S0.2	810	-		0.2 ISM02.4	PPB	2/15/2021 14:02	Std
	2.5 S2.5	10926	-		2.5 ISM02.4	PPB	2/15/2021 14:04	Std
	5 S5.0	22417	-		5 ISM02.4	PPB	2/15/2021 14:07	Std
	7.5 S7.5	32761	-		7.5 ISM02.4	PPB	2/15/2021 14:09	Std
	10 S10.0	43513	-		10 ISM02.4	PPB	2/15/2021 14:11	Std
ICV	ICV	16665	3.7998 -		ISM02.4	PPB	2/15/2021 14:14	SMPL
ICB	ICB	-41	-0.0285 -		ISM02.4	PPB	2/15/2021 14:16	SMPL
CCV083	CCV083	22323	5.0964 -		ISM02.4	PPB	2/15/2021 14:18	SMPL
CCB083	CCB083	-40	-0.0283 -		ISM02.4	PPB	2/15/2021 14:21	SMPL
PB134642BL	PBS642	75	-0.002 -		ISM02.4	PPB	2/15/2021 14:23	SMPL
M1416-02	MBG0T9	616	0.122 -		ISM02.4	PPB	2/15/2021 14:25	SMPL
M1416-03	MBG0T9D	513	0.0984 -		ISM02.4	PPB	2/15/2021 14:30	SMPL
M1416-04	MBG0T9S	11018	2.5058 -		ISM02.4	PPB	2/15/2021 14:32	SMPL
M1416-05	MBG0W0	394	0.0711 -		ISM02.4	PPB	2/15/2021 14:35	SMPL
M1416-06	MBG0W1	477	0.0902 -		ISM02.4	PPB	2/15/2021 14:37	SMPL
M1416-07	MBG0W2	1443	0.3115 -		ISM02.4	PPB	2/15/2021 14:39	SMPL
M1416-08	MBG0W3	285	0.0462 -		ISM02.4	PPB	2/15/2021 14:42	SMPL
CCV084	CCV084	21537	4.9163 -		ISM02.4	PPB	2/15/2021 14:44	SMPL
CCB084	CCB084	-44	-0.0292 -		ISM02.4	PPB	2/15/2021 14:46	SMPL