


LB113141

7471B

INSTRUMENT ID: CV1

 μ Abs.:

4777E



A= 0.0000e+000

B= 2.0772e-004

C= -1.0164e-002

Rho= 0.9998356

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	
0.0	0.000	0.009	0.009	94	0.000	94					
0.2	0.200	0.203	0.003	1028	0.0 %	1028					
2.5	2.500	2.438	-0.062	11788	0.0 %	11788					
5.0	5.000	5.010	0.010	24169	0.0 %	24169					
7.5	7.500	7.625	0.125	36755	0.0 %	36755					
10.0	10.000	9.914	-0.086	47778	0.0 %	47778					

1/05
 -
 2
 -2
 0
 2
 -1

LB113141 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 15 Feb 2021 15:36:40

Sample ID	Extended ID	μ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	SO	94	-	0.00007471B	PPB	15 Feb 2021 15:37:48	S	Std
0.2 - 1	SO.2	1028	-	0.20007471B	PPB	15 Feb 2021 15:40:05	S	Std
2.5 - 1	SO.5	11788	-	2.50007471B	PPB	15 Feb 2021 15:42:22	S	Std
5.0 - 1	SO.5	24169	-	5.00007471B	PPB	15 Feb 2021 15:44:39	S	Std
7.5 - 1	SO.5	36755	-	7.50007471B	PPB	15 Feb 2021 15:46:55	S	Std
10.0 - 1	SO	47778	-	10.00007471B	PPB	15 Feb 2021 15:49:12	S	Std
ICV04 - 1	ICV04	18360	3.8035	-7471B	PPB	15 Feb 2021 15:52:01	U	SMPL
ICB04 - 1	ICB04	31	-0.0037	-7471B	PPB	15 Feb 2021 15:54:16	U	SMPL
CCV85 - 1	CCV85	23986	4.9722	-7471B	PPB	15 Feb 2021 15:56:31	U	SMPL
CCB85 - 1	CCB85	2	-0.0097	-7471B	PPB	15 Feb 2021 15:58:46	U	SMPL
CRA - 1	CRA	989	0.1953	-7471B	PPB	15 Feb 2021 16:01:02	U	SMPL
HighStd - 1	HighStd	47309	9.8168	-7471B	PPB	15 Feb 2021 16:03:18	U	SMPL
ChkStd - 1	ChkStd	33752	7.0007	-7471B	PPB	15 Feb 2021 16:05:33	U	SMPL
PB134660BL - 1	PBS	-50	-0.0205	-7471B	PPB	15 Feb 2021 16:07:50	U	SMPL
PB134660BS - 1	LCSS	18157	3.7614	-7471B	PPB	15 Feb 2021 16:10:08	U	SMPL
M1400-01 - 1	BIN-2	6952	1.4339	-7471B	PPB	15 Feb 2021 16:12:24	U	SMPL
M1402-01 - 1	RO-N9994	1331	0.2663	-7471B	PPB	15 Feb 2021 16:14:41	U	SMPL
M1402-03 - 1	RO-RBR-200058	1413	0.2833	-7471B	PPB	15 Feb 2021 16:16:59	U	SMPL
M1403-01 - 1	72-11939	6744	1.3907	-7471B	PPB	15 Feb 2021 16:38:16	U	SMPL
M1403-01DUP - 1	72-11939DUP	6483	1.3365	-7471B	PPB	15 Feb 2021 16:40:30	U	SMPL
CCV86 - 1	CCV86	25797	5.3483	-7471B	PPB	15 Feb 2021 16:42:46	U	SMPL
CCB86 - 1	CCB86	47	-0.0004	-7471B	PPB	15 Feb 2021 16:45:01	U	SMPL
M1403-01MS - 1	72-11939MS	27789	5.7621	-7471B	PPB	15 Feb 2021 16:47:16	U	SMPL
M1403-01MSD - 1	72-11939MSD	28629	5.9366	-7471B	PPB	15 Feb 2021 16:49:32	U	SMPL
M1403-01LX5 - 1		1286	0.2570	-7471B	PPB	15 Feb 2021 16:51:48	U	SMPL
M1403-01A - 1		28124	5.8317	-7471B	PPB	15 Feb 2021 16:54:04	U	SMPL
CCV87 - 1	CCV87	23639	4.9001	-7471B	PPB	15 Feb 2021 17:03:07	U	SMPL
CCB87 - 1	CCB87	13	-0.0075	-7471B	PPB	15 Feb 2021 17:05:24	U	SMPL