



PERCENT SOLID

Supervisor: apatel
Analyst: Alexander
Date: 2/16/2021

OVENTEMP IN Celsius(°C): 107
Time IN: 17:20
In Date: 02/15/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 105
Time OUT: 09:20
Out Date: 02/16/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB113146

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
M1379-01	DBGP4	1	1.00	2.00	2.00	100.0	P.E.SAMPLE
M1379-02	DBGP5	2	1.00	2.00	2.00	100.0	P.E.SAMPLE
M1379-03	DBGP7	3	1.00	2.00	2.00	100.0	P.E.SAMPLE
M1379-04	DBGY4	4	1.18	9.51	7.48	75.6	
M1379-05	DBGY5	5	1.17	9.84	7.98	78.5	
M1379-06	DBGY6	6	1.16	9.55	8.34	85.6	
M1379-07	DBGY7	7	1.16	9.55	8.42	86.5	
M1379-08	DBGY8	8	1.17	9.6	8.33	84.9	
M1379-09	DBGY9	9	1.17	9.68	8.3	83.8	
M1379-10	DBGZ0	10	1.17	9.7	8.00	80.1	
M1379-11	DBGZ1	11	1.17	9.94	8.74	86.3	
M1379-12	DBGZ2	12	1.16	9.98	8.81	86.7	
M1379-13	DBGZ3	13	1.16	9.9	8.39	82.7	
M1379-14	DBGZ4	14	1.17	9.91	8.58	84.8	
M1379-15	DBGZ5	15	1.14	9.94	8.59	84.7	
M1379-16	DBH00	16	1.16	9.98	8.32	81.2	
M1379-17	DBH01	17	1.15	9.7	8.58	86.9	
M1379-18	DBH15	18	1.16	9.76	8.01	79.7	
M1379-19	DBH16	19	1.15	9.79	7.98	79.1	
M1379-20	DBH16MS	20	1.15	9.79	7.98	79.1	
M1379-21	DBH16MSD	21	1.15	9.79	7.98	79.1	
M1379-22	DBH17	22	1.15	9.55	8.04	82.0	
M1379-23	DBH18	23	1.15	9.57	8.17	83.4	
M1379-24	DBH19	24	1.15	9.78	8.3	82.9	
M1379-25	DBH20	25	1.15	9.77	8.76	88.3	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$