


LB113149

7470A

INSTRUMENT ID: CV2

 μ Abs.:

42373



Linear

A= 0.0000e+000

B= 2.3389e-004

C= -3.4305e-002

Rho= 0.9997511

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.001	-0.001	141	0.000	141					-
0.2	0.200	0.179	-0.021	912	0.0 %	912					-1
2.5	2.500	2.460	-0.040	10665	0.0 %	10665					-2
5.0	5.000	5.038	0.038	21685	0.0 %	21685					1
7.5	7.500	7.648	0.148	32846	0.0 %	32846					2
10.0	10.000	9.876	-0.124	42373	0.0 %	42373					-1

LB113149 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 16 Feb 2021 14:12:16					
Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	141	-	0.0000	7470A	PPB	16 Feb 2021 14:16:02	S	Std
0.2 - 1	50.2	912	-	0.2000	7470A	PPB	16 Feb 2021 14:18:19	S	Std
2.5 - 1	52.5	10665	-	2.5000	7470A	PPB	16 Feb 2021 14:20:37	S	Std
5.0 - 1	55	21685	-	5.0000	7470A	PPB	16 Feb 2021 14:22:54	S	Std
7.5 - 1	57.5	32846	-	7.5000	7470A	PPB	16 Feb 2021 14:25:10	S	Std
10.0 - 1	510	42373	-	10.0000	7470A	PPB	16 Feb 2021 14:27:26	S	Std
ICV05 - 1	ICV05	18081	4.1947	-	7470A	PPB	16 Feb 2021 14:32:37	U	SMPL
ICB05 - 1	ICB05	76	-0.0165	-	7470A	PPB	16 Feb 2021 14:34:53	U	SMPL
CCV88 - 1	CCV88	21982	5.1071	-	7470A	PPB	16 Feb 2021 14:37:08	U	SMPL
CCB88 - 1	CCB88	92	-0.0128	-	7470A	PPB	16 Feb 2021 14:39:24	U	SMPL
CRA - 1	CRA	907	0.1778	-	7470A	PPB	16 Feb 2021 14:41:39	U	SMPL
HighStd - 1	HighStd	41108	9.5805	-	7470A	PPB	16 Feb 2021 14:43:55	U	SMPL
ChkStd - 1	ChkStd	29079	6.7670	-	7470A	PPB	16 Feb 2021 14:46:10	U	SMPL
PB134670BL - 1	PBW	42	-0.0245	-	7470A	PPB	16 Feb 2021 14:48:27	U	SMPL
PB134670BS - 1	LCSW	16632	3.8558	-	7470A	PPB	16 Feb 2021 14:50:43	U	SMPL
M1189-23 - 1	41108	575	0.1002	-	7470A	PPB	16 Feb 2021 14:57:45	U	SMPL
M1189-23DUP - 1	41108DUP	660	0.1201	-	7470A	PPB	16 Feb 2021 15:00:02	U	SMPL
M1189-23MS - 1	41108MS	16369	3.7943	-	7470A	PPB	16 Feb 2021 15:04:52	U	SMPL
M1189-23MSD - 1	41108MSD	15072	3.4909	-	7470A	PPB	16 Feb 2021 15:07:08	U	SMPL
PB134671BL - 1	PBW	-19	-0.0387	-	7470A	PPB	16 Feb 2021 15:09:24	U	SMPL
CCV89 - 1	CCV89	21781	5.0601	-	7470A	PPB	16 Feb 2021 15:11:39	U	SMPL
CCB89 - 1	CCB89	66	-0.0189	-	7470A	PPB	16 Feb 2021 15:13:54	U	SMPL
PB134671BS - 1	LCSW	15227	3.5272	-	7470A	PPB	16 Feb 2021 15:16:10	U	SMPL
M1404-01 - 1	NWB-1579-1582	514	0.0859	-	7470A	PPB	16 Feb 2021 15:18:25	U	SMPL
M1404-01DUP - 1	NWB-1579-1582DUP	387	0.0562	-	7470A	PPB	16 Feb 2021 15:20:41	U	SMPL
M1404-01MS - 1	NWB-1579-1582MS	14946	3.4614	-	7470A	PPB	16 Feb 2021 15:22:57	U	SMPL
M1404-01MSD - 1	NWB-1579-1582MSD	14890	3.4483	-	7470A	PPB	16 Feb 2021 15:25:13	U	SMPL
PB134670TB - 1	PB134655TB	62	-0.0198	-	7470A	PPB	16 Feb 2021 15:27:30	U	SMPL
M1189-23LX5 - 1		193	0.0108	-	7470A	PPB	16 Feb 2021 15:29:47	U	SMPL
M1189-23A - 1		16295	3.7770	-	7470A	PPB	16 Feb 2021 15:32:04	U	SMPL
M1404-01LX5 - 1		131	-0.0037	-	7470A	PPB	16 Feb 2021 15:34:22	U	SMPL
M1404-01A - 1		15115	3.5010	-	7470A	PPB	16 Feb 2021 15:36:37	U	SMPL
CCV90 - 1	CCV90	21580	5.0131	-	7470A	PPB	16 Feb 2021 15:38:54	U	SMPL
CCB90 - 1	CCB90	78	-0.0161	-	7470A	PPB	16 Feb 2021 15:41:09	U	SMPL