



PERCENT SOLID

Supervisor: apatel
Analyst: Alexander
Date: 2/17/2021

OVENTEMP IN Celsius(°C): 107
Time IN: 16:20
In Date: 02/16/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 105
Time OUT: 09:50
Out Date: 02/17/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB113151

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
M1414-01	DBGT3	1	1.18	9.57	7.62	76.8	
M1414-02	DBGT7	2	1.17	9.91	7.31	70.3	
M1414-03	DBGT7MS	3	1.17	9.91	7.31	70.3	
M1414-04	DBGT7MSD	4	1.17	9.91	7.31	70.3	
M1414-05	DBGX1	5	1.17	9.78	8.98	90.7	
M1414-06	DBGX2	6	1.18	9.85	8.56	85.1	
M1414-07	DBGX3	7	1.17	9.98	5.42	48.2	
M1414-08	DBGX4	8	1.17	9.57	7.52	75.6	
M1414-09	DBGX5	9	1.18	9.51	7.7	78.3	
M1414-10	DBGX6	10	1.17	9.94	8.6	84.7	
M1414-11	DBGX7	11	1.17	9.98	9.96	99.8	
M1414-12	DBGZ3	12	1.18	9.88	8.27	81.5	
M1414-13	DBGZ5	13	1.17	9.72	8.56	86.4	
M1414-14	DBGZ9	14	1.17	9.65	8.34	84.6	
M1414-15	DBH01	15	1.18	9.63	8.39	85.3	
M1414-16	DBH08	16	1.18	9.83	8.37	83.1	
M1414-17	DBH09	17	1.18	9.92	8.46	83.3	
M1414-18	DBH10	18	1.17	9.91	8.54	84.3	
M1414-19	DBH11	19	1.14	9.86	7.94	78.0	
M1414-20	DBH23	20	1.14	9.65	8.02	80.8	
M1414-21	DBH27	21	1.14	9.69	8.37	84.6	
M1414-22	DBH34	22	1.18	9.55	9.53	99.8	
M1414-23	VHBLK01	23	1.00	2.00	2.00	100.0	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$