



Analytical Summary Report

Analysis Method: 9040C

Analyst By : Alexander

Parameter: pH

Supervisor Review By : apatel

Run Number: LB113156

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2633
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.4	4.01	02/16/2021	16:10
2	CAL2	1	Water	NA	NA	20.4	7.01	02/16/2021	16:11
3	CAL3	1	Water	NA	NA	20.4	10.03	02/16/2021	16:14
4	ICV	1	Water	NA	NA	20.4	6.97	02/16/2021	16:16
5	CCV1	1	Water	NA	NA	20.3	2.01	02/16/2021	16:40
6	M1425-07	1	Water	NA	NA	18.8	6.85	02/16/2021	16:42
7	M1425-07DUP	1	Water	NA	NA	18.8	6.88	02/16/2021	16:45
8	M1425-08	1	Water	NA	NA	19.0	6.63	02/16/2021	16:47
9	M1425-09	1	Water	NA	NA	18.6	7.98	02/16/2021	16:49
10	M1425-10	1	Water	NA	NA	18.7	6.64	02/16/2021	16:52
11	M1425-11	1	Water	NA	NA	19.3	6.46	02/16/2021	16:55
12	M1425-13	1	Water	NA	NA	19.5	6.70	02/16/2021	16:57
13	CCV2	1	Water	NA	NA	20.3	12.01	02/16/2021	16:59

LB113156
WORKLIST(Hardcopy Internal Chain)

WorkList Name : pH-021621-water

WorkList ID : 146439

Department : Wet-Chemistry

Date : 02-16-2021 15:34:15

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/16/2021	Water	M1425-07	pH	Cool 4 deg C	PSEG03	H22	41055-W	02/16/2021	9040C
02/16/2021	Water	M1425-08	pH	Cool 4 deg C	PSEG03	H22	41056-W	02/16/2021	9040C
02/16/2021	Water	M1425-09	pH	Cool 4 deg C	PSEG03	H22	41058-W	02/16/2021	9040C
02/16/2021	Water	M1425-10	pH	Cool 4 deg C	PSEG03	H22	41059-W	02/16/2021	9040C
02/16/2021	Water	M1425-11	pH	Cool 4 deg C	PSEG03	H22	41060-W	02/16/2021	9040C
02/16/2021	Water	M1425-13	pH	Cool 4 deg C	PSEG03	H22	41061	02/16/2021	9040C

Date/Time 02/16/2021 16:05

Received by: AML(wet-chem)

Relinquished by: KS (Wetchem)

Date/Time 02/16/2021 17:20

Received by: KS (Wetchem)

Relinquished by: AML(wet-chem)