



Analytical Summary Report

Analysis Method: 9040C

Analyst By : Alexander

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB113163

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2633
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.4	4.01	02/17/2021	10:05
2	CAL2	1	Water	NA	NA	20.4	7.01	02/17/2021	10:06
3	CAL3	1	Water	NA	NA	20.4	10.03	02/17/2021	10:08
4	ICV	1	Water	NA	NA	20.4	7.02	02/17/2021	10:11
5	CCV1	1	Water	NA	NA	20.5	2.01	02/17/2021	10:13
6	M1431-01	1	Water	NA	NA	18.8	11.48	02/17/2021	10:15
7	M1431-01DUP	1	Water	NA	NA	18.8	11.47	02/17/2021	10:18
8	CCV2	1	Water	NA	NA	20.5	12.02	02/17/2021	10:20

28113163
WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity-21721 WorkList ID : 146464 Department : Wet-Chemistry Date : 02-17-2021 09:51:14

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/16/2021	Water	M1431-01	Corrosivity	Cool 4 deg C	JACO05	G51	IDW-BR2-AQ-BR-021621	02/16/2021	9040C

Date/Time 02/17/2021 10:00
Received by: AMLwetchem
Relinquished by: KS (Welch)

Date/Time 02/17/2021 10:40
Received by: KS (Welch)
Relinquished by: AMLwetchem