



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB113170

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: FLASHPONT

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.00	80.6	80.9	02/17/2021	09:20
2	M1425-07		1	9	100.00	>212.0	>212.0	02/17/2021	09:45
3	M1425-07DUP		1	9	100.00	>212.0	>212.0	02/17/2021	10:10
4	M1425-08		1	11	100.00	>212.0	>212.0	02/17/2021	10:40
5	M1425-09		1	10	100.00	>212.0	>212.0	02/17/2021	11:15
6	M1425-10		1	9	100.00	>212.0	>212.0	02/17/2021	11:45
7	M1425-11		1	10	100.00	>212.0	>212.0	02/17/2021	12:10
8	M1425-13		1	11	100.00	>212.0	>212.0	02/17/2021	12:35

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 113170

WorkList Name : FP21721 WorkList ID : 146449 Department : Wet-Chemistry Date : 02-17-2021 08:52:16

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/23/2021	Water	M1425-07	B Flash Point	Cool 4 deg C	PSEG03	H22	41055-W	02/16/2021	1010A
02/23/2021	Water	M1425-08	B Flash Point	Cool 4 deg C	PSEG03	H22	41056-W	02/16/2021	1010A
02/23/2021	Water	M1425-09	B Flash Point	Cool 4 deg C	PSEG03	H22	41058-W	02/16/2021	1010A
02/23/2021	Water	M1425-10	B Flash Point	Cool 4 deg C	PSEG03	H22	41059-W	02/16/2021	1010A
02/23/2021	Water	M1425-11	B Flash Point	Cool 4 deg C	PSEG03	H22	41060-W	02/16/2021	1010A
02/23/2021	Water	M1425-13	E Flash Point	Cool 4 deg C	PSEG03	H22	41061	02/16/2021	1010A

Date/Time 02/17/21 08:55
 Received by: KS (Weber)
 Relinquished by: KS (Weber)

Date/Time 02/17/21 12:40
 Received by: KS (Weber)
 Relinquished by: KS (Weber)