


LB 113172

ISM02.4

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 1.9968e-004 slope

C= 4.0391e-003 y-intercept

Rho= 0.9999869

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.021	0.021	85	0.000	85					
0.05	0.050					312					
0.20	0.200	0.209	0.009	1028	0.0 %	1028					
2.50	2.500	2.476	-0.024	12379	0.0 %	12379					
5.00	5.000	4.983	-0.017	24933	0.0 %	24933					
7.50	7.500	7.487	-0.013	37477	0.0 %	37477					
10.0	10.000	10.024	0.024	50179	0.0 %	50179					

700
115
-1000

LB113172

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	85	-		0 ISM02.4	PPB	2/17/2021 14:27	Std
	0.2 S0.2	1028	-		0.2 ISM02.4	PPB	2/17/2021 14:30	Std
	2.5 S2.5	12379	-		2.5 ISM02.4	PPB	2/17/2021 14:32	Std
	5 S5.0	24933	-		5 ISM02.4	PPB	2/17/2021 14:34	Std
	7.5 S7.5	37477	-		7.5 ISM02.4	PPB	2/17/2021 14:36	Std
	10 S10.0	50179	-		10 ISM02.4	PPB	2/17/2021 14:39	Std
ICV	ICV	19597	3.9172	-	ISM02.4	PPB	2/17/2021 14:41	SMPL
ICB	ICB	-9	0.0022	-	ISM02.4	PPB	2/17/2021 14:44	SMPL
CCV096	CCV096	25469	5.0897	-	ISM02.4	PPB	2/17/2021 14:46	SMPL
CCB096	CCB096	-3	0.0034	-	ISM02.4	PPB	2/17/2021 14:48	SMPL
PB134693BL	PBS693	246	0.0532	-	ISM02.4	PPB	2/17/2021 14:50	SMPL
M1409-03	MDBG T3	211	0.0462	-	ISM02.4	PPB	2/17/2021 14:55	SMPL
M1409-04	MDBG T5	732	0.1502	-	ISM02.4	PPB	2/17/2021 14:57	SMPL
M1409-05	MDBG T7	893	0.1824	-	ISM02.4	PPB	2/17/2021 14:59	SMPL
M1409-06	MDBG T7D	855	0.1748	-	ISM02.4	PPB	2/17/2021 15:02	SMPL
M1409-07	MDBG T7S	13891	2.7778	-	ISM02.4	PPB	2/17/2021 15:04	SMPL
M1409-08	MDBG X1	478	0.0995	-	ISM02.4	PPB	2/17/2021 15:06	SMPL
M1409-09	MDBG X2	1136	0.2309	-	ISM02.4	PPB	2/17/2021 15:09	SMPL
M1409-10	MDBG X3	2195	0.4423	-	ISM02.4	PPB	2/17/2021 15:11	SMPL
M1409-11	MDBG X4	869	0.1776	-	ISM02.4	PPB	2/17/2021 15:13	SMPL
M1409-12	MDBG X5	330	0.0699	-	ISM02.4	PPB	2/17/2021 15:15	SMPL
M1409-13	MDBG X6	470	0.0979	-	ISM02.4	PPB	2/17/2021 15:18	SMPL
M1409-14	MDBG X8	1058	0.2153	-	ISM02.4	PPB	2/17/2021 15:20	SMPL
M1409-15	MDBG X9	1445	0.2926	-	ISM02.4	PPB	2/17/2021 15:22	SMPL
M1409-16	MDBG Y0	628	0.1294	-	ISM02.4	PPB	2/17/2021 15:24	SMPL
M1409-17	MDBG Y1	755	0.1548	-	ISM02.4	PPB	2/17/2021 15:27	SMPL
M1409-18	MDBG Y2	7644	1.5304	-	ISM02.4	PPB	2/17/2021 15:29	SMPL
M1409-19	MDBG Y3	1424	0.2884	-	ISM02.4	PPB	2/17/2021 15:31	SMPL
M1409-20	MDBG Y4	4446	0.8918	-	ISM02.4	PPB	2/17/2021 15:34	SMPL
CCV097	CCV097	25719	5.1396	-	ISM02.4	PPB	2/17/2021 15:36	SMPL
CCB097	CCB097	4	0.0048	-	ISM02.4	PPB	2/17/2021 15:38	SMPL
M1409-21	MDBG Y5	1953	0.394	-	ISM02.4	PPB	2/17/2021 15:40	SMPL
M1409-22	MDBG Y6	1196	0.2429	-	ISM02.4	PPB	2/17/2021 15:43	SMPL
M1409-23	MDBG Y7	2935	0.5901	-	ISM02.4	PPB	2/17/2021 15:45	SMPL
M1409-24	MDBG Y8	2540	0.5112	-	ISM02.4	PPB	2/17/2021 15:47	SMPL
PB134694BL	PBS694	229	0.0498	-	ISM02.4	PPB	2/17/2021 15:49	SMPL
M1410-01	MDBG P9	40921	8.1751	-	ISM02.4	PPB	2/17/2021 15:52	SMPL
M1410-02	MDBG Y9	1946	0.3926	-	ISM02.4	PPB	2/17/2021 15:54	SMPL
M1410-03	MDBG Z0	2710	0.5452	-	ISM02.4	PPB	2/17/2021 15:56	SMPL
M1410-04	MDBG Z1	24728	4.9417	-	ISM02.4	PPB	2/17/2021 15:59	SMPL
M1410-05	MDBG Z2	55604	11.107	-	ISM02.4	PPB	2/17/2021 16:01	SMPL
M1410-06	MDBG Z3	409	0.0857	-	ISM02.4	PPB	2/17/2021 16:03	SMPL
M1410-07	MDBG Z4	779	0.1596	-	ISM02.4	PPB	2/17/2021 16:05	SMPL
M1410-08	MDBG Z5	1754	0.3543	-	ISM02.4	PPB	2/17/2021 16:08	SMPL
M1410-09	MDBG Z6	1710	0.3455	-	ISM02.4	PPB	2/17/2021 16:10	SMPL
M1410-10	MDBG Z7	8534	1.7081	-	ISM02.4	PPB	2/17/2021 16:12	SMPL

LB113172

INSTRUMENT ID : CV1

M1410-11	MDBGZ8	1763	0.3561 -	ISM02.4	PPB	2/17/2021 16:14 SMPL
M1410-12	MDBGZ8D	1638	0.3311 -	ISM02.4	PPB	2/17/2021 16:22 SMPL
M1410-13	MDBGZ8S	15191	3.0374 -	ISM02.4	PPB	2/17/2021 16:24 SMPL
M1410-14	MDBGZ9	2096	0.4226 -	ISM02.4	PPB	2/17/2021 16:26 SMPL
M1410-15	MDBH00	2611	0.5254 -	ISM02.4	PPB	2/17/2021 16:29 SMPL
CCV098	CCV098	25360	5.0679 -	ISM02.4	PPB	2/17/2021 16:31 SMPL
CCB098	CCB098	-30	-0.002 -	ISM02.4	PPB	2/17/2021 16:33 SMPL
M1410-16	MDBH01	1736	0.3507 -	ISM02.4	PPB	2/17/2021 16:36 SMPL
M1410-17	MDBH02	2749	0.553 -	ISM02.4	PPB	2/17/2021 16:38 SMPL
M1410-18	MDBH08	1646	0.3327 -	ISM02.4	PPB	2/17/2021 16:40 SMPL
M1410-19	MDBH09	840	0.1718 -	ISM02.4	PPB	2/17/2021 16:42 SMPL
M1410-20	MDBH10	1247	0.253 -	ISM02.4	PPB	2/17/2021 16:45 SMPL
M1410-21	MDBH11	1215	0.2466 -	ISM02.4	PPB	2/17/2021 16:47 SMPL
M1410-22	MDBH14	1181	0.2399 -	ISM02.4	PPB	2/17/2021 16:49 SMPL
M1410-23	MDBH15	781	0.16 -	ISM02.4	PPB	2/17/2021 16:51 SMPL
PB134695BL	PBS695	242	0.0524 -	ISM02.4	PPB	2/17/2021 16:54 SMPL
M1411-01	MDBH16	301	0.0641 -	ISM02.4	PPB	2/17/2021 16:56 SMPL
M1411-02	MDBH16D	291	0.0621 -	ISM02.4	PPB	2/17/2021 16:58 SMPL
M1411-03	MDBH16S	12966	2.5931 -	ISM02.4	PPB	2/17/2021 17:00 SMPL
M1411-04	MDBH17	304	0.0647 -	ISM02.4	PPB	2/17/2021 17:03 SMPL
M1411-05	MDBH18	287	0.0613 -	ISM02.4	PPB	2/17/2021 17:05 SMPL
M1411-06	MDBH19	264	0.0568 -	ISM02.4	PPB	2/17/2021 17:07 SMPL
M1411-07	MDBH20	354	0.0747 -	ISM02.4	PPB	2/17/2021 17:10 SMPL
M1411-08	MDBH21	1064	0.2165 -	ISM02.4	PPB	2/17/2021 17:12 SMPL
M1411-09	MDBH22	361	0.0761 -	ISM02.4	PPB	2/17/2021 17:14 SMPL
M1411-10	MDBH23	814	0.1666 -	ISM02.4	PPB	2/17/2021 17:16 SMPL
M1411-11	MDBH24	700	0.1438 -	ISM02.4	PPB	2/17/2021 17:19 SMPL
CCV099	CCV099	25487	5.0933 -	ISM02.4	PPB	2/17/2021 17:21 SMPL
CCB099	CCB099	-13	0.0014 -	ISM02.4	PPB	2/17/2021 17:23 SMPL
M1411-12	MDBH25	737	0.1512 -	ISM02.4	PPB	2/17/2021 17:25 SMPL
M1411-13	MDBH26	384	0.0807 -	ISM02.4	PPB	2/17/2021 17:28 SMPL
M1411-14	MDBH27	688	0.1414 -	ISM02.4	PPB	2/17/2021 17:30 SMPL
M1411-15	MDBH35	1942	0.3918 -	ISM02.4	PPB	2/17/2021 17:32 SMPL
M1409-01	MDBGP8	10210	2.0428 -	ISM02.4	PPB	2/17/2021 17:37 SMPL
M1410-05DLX2	MDBGZ2	27803	5.5557 -	ISM02.4	PPB	2/17/2021 17:44 SMPL
CCV001	CCV000	25535	5.1028 -	ISM02.4	PPB	2/17/2021 17:46 SMPL
CCB001	CCB001	-26	-0.0012 -	ISM02.4	PPB	2/17/2021 17:48 SMPL