



Analytical Summary Report

Analysis Method: 9040C

Analyst By : Alexander

Parameter: pH

Supervisor Review By : apatel

Run Number: LB113177

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2633
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.5	4.01	02/17/2021	16:03
2	CAL2	1	Water	NA	NA	20.5	7.01	02/17/2021	16:05
3	CAL3	1	Water	NA	NA	20.5	10.03	02/17/2021	16:06
4	ICV	1	Water	NA	NA	20.5	6.99	02/17/2021	16:09
5	CCV1	1	Water	NA	NA	20.4	2.02	02/17/2021	16:11
6	M1438-01	1	Water	NA	NA	19.4	6.60	02/17/2021	16:13
7	M1438-01DUP	1	Water	NA	NA	19.4	6.59	02/17/2021	16:16
8	M1438-03	1	Water	NA	NA	19.3	6.42	02/17/2021	16:18
9	CCV2	1	Water	NA	NA	20.4	12.02	02/17/2021	16:20

18113177
WORKLIST(Hardcopy-Internal Chain)

WorkList Name : pH-water-21721 WorkList ID : 146467 Department : Wet-Chemistry Date : 02-17-2021 09:51:37

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/17/2021	Water	M1438-01	pH	Cool 4 deg C	PSEG03	H32	TRE-1157	02/17/2021	9040C
02/17/2021	Water	M1438-03	pH	Cool 4 deg C	PSEG03	H32	TRE-1149-56-58	02/17/2021	9040C

Date/Time 02/17/2021 16:00
 Received by: AmLwetcher
 Relinquished by: KS (Wetchem)

Date/Time 02/17/2021 17:20
 Received by: KS (Wetchem)
 Relinquished by: AmLwetcher