


LB 113187

7471B

INSTRUMENT ID: CY1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	
0.0	0.000	0.036	0.036	45	0.000	45					900
0.2	0.200	0.195	-0.005	990	0.0 %	990					13
2.5	2.500	2.415	-0.085	14204	0.0 %	14204					13
5.0	5.000	5.046	0.046	29866	0.0 %	29866					1
7.5	7.500	7.547	0.047	44756	0.0 %	44756					1
10.0	10.000	9.962	-0.038	59133	0.0 %	59133					0

LB113187 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 18 Feb 2021 11:28:41

Sample ID	Extended ID	µ Abs.	Conc.	Strnd Conc	Method	Units	Date	Type	Type
0.0 - 1	50	45	-	0.0000	7471B	PPB	18 Feb 2021 11:30:04	S	Std
0.2 - 1	50.2	990	-	0.2000	7471B	PPB	18 Feb 2021 11:32:22	S	Std
2.5 - 1	52.5	14204	-	2.5000	7471B	PPB	18 Feb 2021 11:34:40	S	Std
5.0 - 1	55	29866	-	5.0000	7471B	PPB	18 Feb 2021 11:36:57	S	Std
7.5 - 1	57.5	44756	-	7.5000	7471B	PPB	18 Feb 2021 11:41:51	S	Std
10.0 - 1	57.8	59133	-	10.0000	7471B	PPB	18 Feb 2021 11:46:38	S	Std
ICV07 - 1	ICV07	22531	3.8134	-	7471B	PPB	18 Feb 2021 11:49:17	U	SMPL
ICB07 - 1	ICB07	28	0.0332	-	7471B	PPB	18 Feb 2021 11:53:41	U	SMPL
CCV02 - 1	CCV02	30220	5.1051	-	7471B	PPB	18 Feb 2021 11:55:56	U	SMPL
CCB02 - 1	CCB02	-54	0.0194	-	7471B	PPB	18 Feb 2021 11:58:11	U	SMPL
CRA - 1	CRA	980	0.1931	-	7471B	PPB	18 Feb 2021 12:00:26	U	SMPL
HighStd - 1	HighStd	59765	10.0682	-	7471B	PPB	18 Feb 2021 12:02:41	U	SMPL
ChkStd - 1	ChkStd	41957	7.0767	-	7471B	PPB	18 Feb 2021 12:04:56	U	SMPL
PB134709BL - 1	PBS	-71	0.0166	-	7471B	PPB	18 Feb 2021 12:07:14	U	SMPL
PB134709BS - 1	LCSS	21198	3.5895	-	7471B	PPB	18 Feb 2021 12:09:31	U	SMPL
M1422-01 - 1	BU-03-021621	820	0.1663	-	7471B	PPB	18 Feb 2021 12:11:47	U	SMPL
M1423-01 - 1	TR-05-021621	1926	0.3520	-	7471B	PPB	18 Feb 2021 12:14:04	U	SMPL
M1424-01 - 1	OR-02-021621	1370	0.2586	-	7471B	PPB	18 Feb 2021 12:16:21	U	SMPL
M1430-08 - 1	WO-1	8795	1.5059	-	7471B	PPB	18 Feb 2021 12:19:26	U	SMPL
M1430-08DUP - 1	WO-1DUP	9319	1.5940	-	7471B	PPB	18 Feb 2021 12:21:42	U	SMPL
CCV03 - 1	CCV03	30053	5.0770	-	7471B	PPB	18 Feb 2021 12:23:57	U	SMPL
CCB03 - 1	CCB03	-37	0.0223	-	7471B	PPB	18 Feb 2021 12:26:12	U	SMPL
M1430-08MS - 1	WO-1MS	33324	5.6265	-	7471B	PPB	18 Feb 2021 12:28:28	U	SMPL
M1430-08MSD - 1	WO-1MSD	33282	5.6194	-	7471B	PPB	18 Feb 2021 12:30:44	U	SMPL
M1434-02 - 1	SAND-SOIL-PILE	3802	0.6672	-	7471B	PPB	18 Feb 2021 12:33:00	U	SMPL
M1435-01 - 1	SU-04-021721	2687	0.4799	-	7471B	PPB	18 Feb 2021 12:35:16	U	SMPL
M1439-01 - 1	SP-142-150	37645	6.3524	-	7471B	PPB	18 Feb 2021 12:37:32	U	SMPL
M1440-01 - 1	SP-227-233-A	22328	3.7793	-	7471B	PPB	18 Feb 2021 12:39:48	U	SMPL
M1440-02 - 1	SP-227-233-B	24526	4.1485	-	7471B	PPB	18 Feb 2021 12:42:06	U	SMPL
M1441-01 - 1	SP-234-A	13938	2.3699	-	7471B	PPB	18 Feb 2021 12:44:24	U	SMPL
M1441-02 - 1	SP-234-B	9831	1.6800	-	7471B	PPB	18 Feb 2021 12:46:41	U	SMPL
M1446-01 - 1	R-LA-3-5-WC	1359	0.2568	-	7471B	PPB	18 Feb 2021 12:48:57	U	SMPL
CCV04 - 1	CCV04	30143	5.0921	-	7471B	PPB	18 Feb 2021 12:51:13	U	SMPL
CCB04 - 1	CCB04	-39	0.0220	-	7471B	PPB	18 Feb 2021 12:53:29	U	SMPL
M1430-08LX5 - 1		1759	0.3240	-	7471B	PPB	18 Feb 2021 12:55:45	U	SMPL
M1430-08A - 1		33093	5.5877	-	7471B	PPB	18 Feb 2021 12:58:01	U	SMPL
CCV05 - 1	CCV05	30067	5.0794	-	7471B	PPB	18 Feb 2021 13:00:17	U	SMPL
CCB05 - 1	CCB05	-36	0.0225	-	7471B	PPB	18 Feb 2021 13:02:33	U	SMPL