

LB113188


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

41402



A= 0.0000e+000

B= 2.4275e-004

C= 1.0563e-002

Rho= 0.9996745

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.019	0.019	35	0.000	35					
0.2	0.200	0.160	-0.040	616	0.0 %	616					
2.5	2.500	2.506	0.006	10281	0.0 %	10281					
5.0	5.000	5.129	0.129	21085	0.0 %	21085					
7.5	7.500	7.324	-0.176	30129	0.0 %	30129					
10.0	10.000	10.061	0.061	41402	0.0 %	41402					

20
 0
 3
 2
 1

LB113188 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 18 Feb 2021 14:03:19					
Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	S0	35	-	0.0000	7470A	PPB	18 Feb 2021 14:03:30	S	Std
0.2 - 1	S1.2	616	-	0.2000	7470A	PPB	18 Feb 2021 14:05:49	S	Std
2.5 - 1	S2.5	10281	-	2.5000	7470A	PPB	18 Feb 2021 14:08:06	S	Std
5.0 - 1	S5	21085	-	5.0000	7470A	PPB	18 Feb 2021 14:10:23	S	Std
7.5 - 1	S7.5	30129	-	7.5000	7470A	PPB	18 Feb 2021 14:12:39	S	Std
10.0 - 1	S10	41402	-	10.0000	7470A	PPB	18 Feb 2021 14:14:55	S	Std
ICV08 - 1	ICV08	16380	3.9869	-	7470A	PPB	18 Feb 2021 14:17:44	U	SMPL
ICB08 - 1	ICB08	28	0.0174	-	7470A	PPB	18 Feb 2021 14:20:34	U	SMPL
CCV06 - 1	CCV06	21331	5.1887	-	7470A	PPB	18 Feb 2021 14:22:49	U	SMPL
CCB06 - 1	CCB06	-31	0.0030	-	7470A	PPB	18 Feb 2021 14:25:04	U	SMPL
CRA - 1	CRA	605	0.1574	-	7470A	PPB	18 Feb 2021 14:27:19	U	SMPL
HighStd - 1	HighStd	41273	10.0297	-	7470A	PPB	18 Feb 2021 14:29:35	U	SMPL
ChkStd - 1	ChkStd	26944	6.5513	-	7470A	PPB	18 Feb 2021 14:32:24	U	SMPL
PB134726BL - 1	PBW	-46	-0.0006	-	7470A	PPB	18 Feb 2021 14:34:40	U	SMPL
PB134726BS - 1	LCSW	14887	3.6244	-	7470A	PPB	18 Feb 2021 14:36:56	U	SMPL
M1434-03 - 1	SAND-SOIL-PILE	78	0.0295	-	7470A	PPB	18 Feb 2021 14:39:12	U	SMPL
M1434-03DUP - 1	SAND-SOIL-PILEDUP	103	0.0356	-	7470A	PPB	18 Feb 2021 14:41:28	U	SMPL
M1434-03MS - 1	SAND-SOIL-PILEMS	15150	3.6883	-	7470A	PPB	18 Feb 2021 14:43:45	U	SMPL
M1434-03MSD - 1	SAND-SOIL-PILEMSD	15182	3.6960	-	7470A	PPB	18 Feb 2021 14:46:03	U	SMPL
PB134727BL - 1	PBW	18	0.0149	-	7470A	PPB	18 Feb 2021 14:49:01	U	SMPL
CCV07 - 1	CCV07	21019	5.1130	-	7470A	PPB	18 Feb 2021 14:51:16	U	SMPL
CCB07 - 1	CCB07	-40	0.0009	-	7470A	PPB	18 Feb 2021 14:53:32	U	SMPL
PB134727BS - 1	LCSW	14465	3.5220	-	7470A	PPB	18 Feb 2021 14:55:47	U	SMPL
M1438-01 - 1	TRE-1157	5	0.0118	-	7470A	PPB	18 Feb 2021 14:58:03	U	SMPL
M1438-01DUP - 1	TRE-1157DUP	37	0.0195	-	7470A	PPB	18 Feb 2021 15:00:19	U	SMPL
M1438-01MS - 1	TRE-1157MS	10185	2.4830	-	7470A	PPB	18 Feb 2021 15:02:34	U	SMPL
M1438-01MSD - 1	TRE-1157MSD	10098	2.4619	-	7470A	PPB	18 Feb 2021 15:04:51	U	SMPL
M1438-03 - 1	TRE-1149-56-58	-65	-0.0052	-	7470A	PPB	18 Feb 2021 15:07:07	U	SMPL
CCV08 - 1	CCV08	20862	5.0749	-	7470A	PPB	18 Feb 2021 15:09:24	U	SMPL
CCB08 - 1	CCB08	-54	-0.0025	-	7470A	PPB	18 Feb 2021 15:11:41	U	SMPL
PB134726TB - 1	PB134710TB	5	0.0118	-	7470A	PPB	18 Feb 2021 15:13:58	U	SMPL
PB134727TB - 1	PB134711TB	8	0.0125	-	7470A	PPB	18 Feb 2021 15:16:14	U	SMPL
M1434-03LX5 - 1		33	0.0186	-	7470A	PPB	18 Feb 2021 15:18:29	U	SMPL
M1434-03A - 1		15987	3.8915	-	7470A	PPB	18 Feb 2021 15:20:45	U	SMPL
M1438-01LX5 - 1		-24	0.0047	-	7470A	PPB	18 Feb 2021 15:23:01	U	SMPL
M1438-01A - 1		16149	3.9308	-	7470A	PPB	18 Feb 2021 15:25:18	U	SMPL
CCV09 - 1	CCV09	20864	5.0754	-	7470A	PPB	18 Feb 2021 15:27:33	U	SMPL
CCB09 - 1	CCB09C	-30	0.0033	-	7470A	PPB	18 Feb 2021 15:29:49	U	SMPL