



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB113195

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.3	02/19/2021	10:32
2	M1431-01		1	9	100.00	>212.0	>212.0	02/19/2021	11:05
3	M1438-01		1	9	100.00	>212.0	>212.0	02/19/2021	11:40
4	M1438-03		1	11	100.00	>212.0	>212.0	02/19/2021	12:05
5	M1438-03DUP		1	11	100.00	>212.0	>212.0	02/19/2021	12:35

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 113195

WorkList Name : fp21921

WorkList ID : 146528

Department : Wet-Chemistry

Date : 02-19-2021 09:59:42

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/23/2021	Water	M1431-01	Flash Point	Cool 4 deg C	JACO05	G51	IDW-BR2-AQ-BR-021621	02/16/2021	1010A
02/22/2021	Water	M1438-01	Flash Point	Cool 4 deg C	PSEG03	H32	TRE-1157	02/17/2021	1010A
03/02/2021	Water	M1438-03	Flash Point	Cool 4 deg C	PSEG03	H32	TRE-1149-56-58	02/17/2021	1010A

Date/Time 02/19/21 10:00
 Received by: KS (Wetchem)
 Relinquished by: PC CExt (ab)

Date/Time 02/19/21 12:40
 Received by: RC CExt (ab)
 Relinquished by: KS (Wetchem)