


LB113196

SFAM01.1

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 2.4245e-004 *slope*C= -3.9522e-002 *y-intercept*

Rho= 0.9999393

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%
0.00	0.000	-0.028	-0.028	46	0.000	46					-
0.05	0.050	0.045	-0.005	350	0.0 %	350					-10
0.20	0.200	0.193	-0.007	957	0.0 %	957					-4
2.50	2.500	2.500	0.000	10476	0.0 %	10476					0
5.00	5.000	5.055	0.055	21014	0.0 %	21014					1
7.50	7.500	7.554	0.054	31321	0.0 %	31321					1
10.0	10.000	9.931	-0.069	41122	0.0 %	41122					-1

LB113196
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	46	-		0 SFAM01.1	PPB	2/19/2021 13:58	Std
	0.05 S01	350	-	0.05	SFAM01.1	PPB	2/19/2021 14:00	Std
	0.2 S02	957	-	0.2	SFAM01.1	PPB	2/19/2021 14:05	Std
	2.5 S03	10476	-	2.5	SFAM01.1	PPB	2/19/2021 14:07	Std
	5 S04	21014	-	5	SFAM01.1	PPB	2/19/2021 14:10	Std
	7.5 S05	31321	-	7.5	SFAM01.1	PPB	2/19/2021 14:12	Std
	10 S06	41122	-	10	SFAM01.1	PPB	2/19/2021 14:17	Std
ICV013	ICV013	16217	3.8923	-	SFAM01.1	PPB	2/19/2021 14:20	SMPL
ICB013	ICB013	68	-0.023	-	SFAM01.1	PPB	2/19/2021 14:32	SMPL
CCV010	CCV010	21381	5.1443	-	SFAM01.1	PPB	2/19/2021 14:34	SMPL
CCB010	CCB010	146	-0.0041	-	SFAM01.1	PPB	2/19/2021 14:41	SMPL
PB134745BL	PBW745	104	-0.0143	-	SFAM01.1	PPB	2/19/2021 14:44	SMPL
M1344-01	MDL-WATER-QT2-2021-03	365	0.049	-	SFAM01.1	PPB	2/19/2021 14:46	SMPL
CCV011	CCV011	21039	5.0614	-	SFAM01.1	PPB	2/19/2021 14:48	SMPL
CCB011	CCB011	194	0.0075	-	SFAM01.1	PPB	2/19/2021 14:53	SMPL