

Analysis Method: 9034
 Parameter: Reactive Sulfide
 Run Number: LB113201

ANALYST: ketankumar
 SUPERVISOR REVIEW BY: apatel
 Constant: 16000
 Normality1: 0.025
 Normality2: 0.025

Reagent/Standard	Lot/Log #
Sodium thiosulfate titrant 0.025N	WP88904
IODINE SOLUTION .025N 1L	W2763
Starch Solution, 4L	W2703

Seq	Lab ID	True Value (mg/l)	DF	Initial Weight (g)	Final Volume (ml)	T1 (ml)	T2 Initial	T2 Final	T2 Diff. (ml)	T1 - T2 Diff (mL)	Value Corrected With Blank	Result (ppm)	Anal Date	Anal Time
1	PB134737BL		1	5.00	50	2.00	0.00	1.90	1.90	0.10	0.00	0.00	02/19/2021	14:05
2	M1434-02		1	5.02	50	2.00	1.90	3.76	1.86	0.14	0.04	3.19	02/19/2021	14:08
3	M1434-02DUP		1	5.04	50	2.00	3.76	5.64	1.88	0.12	0.02	1.59	02/19/2021	14:12
4	M1436-01		1	5.05	50	2.00	5.64	7.48	1.84	0.16	0.06	4.75	02/19/2021	14:15
5	M1436-03		1	5.04	50	2.00	0.00	1.86	1.86	0.14	0.04	3.17	02/19/2021	14:18
6	M1437-01		1	5.06	50	2.00	1.86	3.70	1.84	0.16	0.06	4.74	02/19/2021	14:22
7	M1437-03		1	5.07	50	2.00	3.70	5.58	1.88	0.12	0.02	1.58	02/19/2021	14:25

T1 = Titrant1

T2 = Titrant2

T2 Diff = T2 Final - T2 Initial

Value Corrected With Blank = ((T1 - T2 Diff) - Blank Correction(BL))

Result = ((T1 * Normality1) - ((T1 - Value Corrected With Blank) * Normality2)) * Constant / Initial Volume