


LB113202

SFAM01.1

INSTRUMENT ID: CV1

Linear



A= 0.0000e+000

B= 2.2622e-004 *slope*

C= -4.3182e-002 *y-intercept*

Rho= 0.9996314

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	% i)
0.00	0.000	-0.021	-0.021	97	0.000	97					-
0.05	0.050	0.042	-0.008	377	0.0 %	377					-16
0.20	0.200	0.194	-0.006	1050	0.0 %	1050					-3
2.50	2.500	2.437	-0.063	10965	0.0 %	10965					-3
5.00	5.000	5.090	0.090	22691	0.0 %	22691					2
7.50	7.500	7.675	0.175	34116	0.0 %	34116					2
10.0	10.000	9.833	-0.167	43655	0.0 %	43655					-2

LB113202
INSTRUMENT ID : 1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	97	-		0 SFAM01.1	PPB	2/19/2021 15:33	Std
	0.05 S01	377	-		0.05 SFAM01.1	PPB	2/19/2021 15:35	Std
	0.2 S02	1050	-		0.2 SFAM01.1	PPB	2/19/2021 15:40	Std
	2.5 S03	10965	-		2.5 SFAM01.1	PPB	2/19/2021 15:42	Std
	5 S04	22691	-		5 SFAM01.1	PPB	2/19/2021 15:44	Std
	7.5 S05	34116	-		7.5 SFAM01.1	PPB	2/19/2021 15:46	Std
	10 S06	43655	-		10 SFAM01.1	PPB	2/19/2021 15:51	Std
ICV014	ICV014	16877	3.7748 -		SFAM01.1	PPB	2/19/2021 15:54	SMPL
ICB014	ICB014	112	-0.0178 -		SFAM01.1	PPB	2/19/2021 15:56	SMPL
CCV012	CCV012	23036	5.1681 -		SFAM01.1	PPB	2/19/2021 16:00	SMPL
CCB012	CCB012	138	-0.012 -		SFAM01.1	PPB	2/19/2021 16:03	SMPL
PB134746BL	PBS746	129	-0.014 -		SFAM01.1	PPB	2/19/2021 16:08	SMPL
M1344-03	MDL-SOIL-QT2-2021-03	426	0.0532 -		SFAM01.1	PPB	2/19/2021 16:15	SMPL
CCV013	CCV013	22683	5.0883 -		SFAM01.1	PPB	2/19/2021 16:17	SMPL
CCB013	CCB013	185	-0.0013 -		SFAM01.1	PPB	2/19/2021 16:19	SMPL