

LB 113208


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

43681



A= 0.0000e+000

B= 2.2712e-004 slope

C= -2.5890e-002 y-intercept

Rho= 0.9998242

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.020	-0.020	28	0.000	28					0
0.05	0.050	0.050	-0.000	332	0.0 %	332					5
0.20	0.200	0.209	0.009	1034	0.0 %	1034					13
2.50	2.500	2.430	-0.070	10815	0.0 %	10815					2
5.00	5.000	5.082	0.082	22488	0.0 %	22488					1
7.50	7.500	7.604	0.104	33596	0.0 %	33596					1
10.0	10.000	9.895	-0.105	43681	0.0 %	43681					

LB113208

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	28	-		0 SFAM01.1	PPB	2/22/2021 13:48	Std
	0.05 S01	332	-		0.05 SFAM01.1	PPB	2/22/2021 13:50	Std
	0.2 S02	1034	-		0.2 SFAM01.1	PPB	2/22/2021 13:52	Std
	2.5 S03	10815	-		2.5 SFAM01.1	PPB	2/22/2021 13:55	Std
	5 S04	22488	-		5 SFAM01.1	PPB	2/22/2021 13:57	Std
	7.5 S05	33596	-		7.5 SFAM01.1	PPB	2/22/2021 13:59	Std
	10 S06	43681	-		10 SFAM01.1	PPB	2/22/2021 14:01	Std
ICV015	ICV015	17024	3.8406	-	SFAM01.1	PPB	2/22/2021 14:04	SMPL
ICB015	ICB015	18	-0.0218	-	SFAM01.1	PPB	2/22/2021 14:09	SMPL
CCV014	CCV014	22427	5.0677	-	SFAM01.1	PPB	2/22/2021 14:11	SMPL
CCB014	CCB014	33	-0.0184	-	SFAM01.1	PPB	2/22/2021 14:16	SMPL
PB134769BL	PBS769	45	-0.0157	-	SFAM01.1	PPB	2/22/2021 14:18	SMPL
M1344-04	MDL-SOIL-QT2-2021-04	360	0.0559	-	SFAM01.1	PPB	2/22/2021 14:23	SMPL
CCV015	CCV015	22682	5.1256	-	SFAM01.1	PPB	2/22/2021 14:25	SMPL
CCB015	CCB015	16	-0.0223	-	SFAM01.1	PPB	2/22/2021 14:30	SMPL