


LB113209

7471B

INSTRUMENT ID: CV1

 μ Abs.:

57722



Linear

A= 0.0000e+000

B= 1.7625e-004

C= -1.4580e-001

Rho= 0.9999180

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.014	-0.014	745	0.000	745					
0.2	0.200	0.169	-0.031	1784	0.0 %	1784					
2.5	2.500	2.579	0.079	15458	0.0 %	15458					
5.0	5.000	5.011	0.011	29257	0.0 %	29257					
7.5	7.500	7.428	-0.072	42974	0.0 %	42974					
10.0	10.000	10.028	0.028	57722	0.0 %	57722					

11
-16
3
0
-1
0

LB113209 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 22 Feb 2021 14:58:52

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	745	-	0.0000	7471B	PPB	22 Feb 2021 15:00:32	S	Std
0.2 - 1	50.2	1784	-	0.2000	7471B	PPB	22 Feb 2021 15:04:36	S	Std
2.5 - 1	52.5	15458	-	2.5000	7471B	PPB	22 Feb 2021 15:06:53	S	Std
5.0 - 1	55	29257	-	5.0000	7471B	PPB	22 Feb 2021 15:09:10	S	Std
7.5 - 1	57.5	42974	-	7.5000	7471B	PPB	22 Feb 2021 15:11:26	S	Std
10.0 - 1	56	57722	-	10.0000	7471B	PPB	22 Feb 2021 15:13:43	S	Std
ICV09 - 1	ICV09	23354	3.9704	-	7471B	PPB	22 Feb 2021 15:16:28	U	SMPL
ICB09 - 1	ICB09	566	-0.0460	-	7471B	PPB	22 Feb 2021 15:21:14	U	SMPL
CCV16 - 1	CCV16	29201	5.0009	-	7471B	PPB	22 Feb 2021 15:23:29	U	SMPL
CCB16 - 1	CCB16	662	-0.0291	-	7471B	PPB	22 Feb 2021 15:25:44	U	SMPL
CRA - 1	CRA	1761	0.1646	-	7471B	PPB	22 Feb 2021 15:28:00	U	SMPL
HighStd - 1	HighStd	58189	10.1102	-	7471B	PPB	22 Feb 2021 15:30:16	U	SMPL
ChkStd - 1	ChkStd	40316	6.9600	-	7471B	PPB	22 Feb 2021 15:35:05	U	SMPL
PB134761BL - 1	PBS	792	-0.0062	-	7471B	PPB	22 Feb 2021 15:37:21	U	SMPL
PB134761BS - 1	LCSS	21142	3.5805	-	7471B	PPB	22 Feb 2021 15:39:37	U	SMPL
M1452-01 - 1	TP-1	54434	9.4483	-	7471B	PPB	22 Feb 2021 15:41:54	U	SMPL
M1452-05 - 1	TP-2	34608	5.9539	-	7471B	PPB	22 Feb 2021 15:44:11	U	SMPL
M1452-09 - 1	TP-3	48511	8.4044	-	7471B	PPB	22 Feb 2021 15:46:29	U	SMPL
M1452-13 - 1	TP-4	28347	4.8504	-	7471B	PPB	22 Feb 2021 15:48:47	U	SMPL
M1452-17 - 1	TP-5	17615	2.9589	-	7471B	PPB	22 Feb 2021 15:51:38	U	SMPL
CCV17 - 1	CCV17	29256	5.0106	-	7471B	PPB	22 Feb 2021 15:54:23	U	SMPL
CCB17 - 1	CCB17	817	-0.0018	-	7471B	PPB	22 Feb 2021 15:59:14	U	SMPL
M1452-21 - 1	TP-6	34415	5.9199	-	7471B	PPB	22 Feb 2021 16:01:29	U	SMPL
M1452-25 - 1	TP-7	18633	3.1383	-	7471B	PPB	22 Feb 2021 16:03:44	U	SMPL
M1452-25DUP - 1	TP-7DUP	16871	2.8278	-	7471B	PPB	22 Feb 2021 16:06:00	U	SMPL
M1452-25MS - 1	TP-7MS	36648	6.3135	-	7471B	PPB	22 Feb 2021 16:08:16	U	SMPL
M1452-25MSD - 1	TP-7MSD	37963	6.5453	-	7471B	PPB	22 Feb 2021 16:10:32	U	SMPL
M1452-25LX5 - 1		4162	0.5878	-	7471B	PPB	22 Feb 2021 16:12:49	U	SMPL
M1452-25A - 1		40979	7.0769	-	7471B	PPB	22 Feb 2021 16:15:37	U	SMPL
CCV18 - 1	CCV18	29397	5.0355	-	7471B	PPB	22 Feb 2021 16:17:54	U	SMPL
CCB18 - 1	CCB18	677	-0.0265	-	7471B	PPB	22 Feb 2021 16:20:12	U	SMPL