



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Alexander

Parameter: pH

Supervisor Review By : apatel

Run Number: LB113218

Slope : 98.9

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2633
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.4	4.01	02/22/2021	16:46
2	CAL2	1	Water	NA	NA	20.5	7.01	02/22/2021	16:47
3	CAL3	1	Water	NA	NA	20.5	10.02	02/22/2021	16:50
4	ICV	1	Water	NA	NA	20.5	6.99	02/22/2021	16:53
5	CCV1	1	Water	NA	NA	20.4	2.02	02/22/2021	16:55
6	M1461-01	1	Solid	20.03	20	18.6	5.53	02/22/2021	16:57
7	M1461-01DUP	1	Solid	20.04	20	18.6	5.51	02/22/2021	17:00
8	CCV2	1	Water	NA	NA	20.4	12.02	02/22/2021	17:02

LB113218

WORKLIST(Hardcopy Internal Chain)

WorkList Name : pH-22221 WorkList ID : 146576 Department : Wet-Chemistry Date : 02-22-2021 13:00:33

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/22/2021	Solid	M1461-01	pH	Cool 4 deg C	PSEG03	I11	31155	02/22/2021	9045D

Date/Time 02/22/2021 16:48
Received by: PM(wet-chem)
Relinquished by: KS (wet-chem)

Date/Time 02/22/2021 17:40
Received by: KS (wet-chem)
Relinquished by: AKC(wet-chem)