


LB113225

SFAM01.1

INSTRUMENT ID: CV1

 μ Abs.:

40300



Linear

A= 0.0000e+000

B= 2.4584e-004 slope

C= -3.2969e-002 y-intercept

Rho= 0.9997828

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.023	-0.023	42	0.000	42					-
0.05	0.050	0.049	-0.001	332	0.0 %	332					-2
0.20	0.200	0.154	-0.046	760	0.0 %	760					-23
2.50	2.500	2.548	0.048	10497	0.0 %	10497					2
5.00	5.000	5.003	0.003	20483	0.0 %	20483					6
7.50	7.500	7.646	0.146	31234	0.0 %	31234					2
10.0	10.000	9.874	-0.126	40300	0.0 %	40300					-1

LB113225
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	42	-		0 SFAM01.1	PPB	2/23/2021 12:47	Std
	0.05 S01	332	-		0.05 SFAM01.1	PPB	2/23/2021 12:49	Std
	0.2 S02	760	-		0.2 SFAM01.1	PPB	2/23/2021 12:57	Std
	2.5 S03	10497	-		2.5 SFAM01.1	PPB	2/23/2021 13:00	Std
	5 S04	20483	-		5 SFAM01.1	PPB	2/23/2021 13:05	Std
	7.5 S05	31234	-		7.5 SFAM01.1	PPB	2/23/2021 13:08	Std
	10 S06	40300	-		10 SFAM01.1	PPB	2/23/2021 13:15	Std
ICV016	ICV016	16146	3.9364	-	SFAM01.1	PPB	2/23/2021 13:19	SMPL
ICB016	ICB016	212	0.0191	-	SFAM01.1	PPB	2/23/2021 13:24	SMPL
CCV019	CCV019	20427	4.9888	-	SFAM01.1	PPB	2/23/2021 13:26	SMPL
CCB019	CCB019	94	-0.0099	-	SFAM01.1	PPB	2/23/2021 13:29	SMPL
PB134786BL	PBW786	181	0.0115	-	SFAM01.1	PPB	2/23/2021 13:31	SMPL
M1344-02	MDL-WATER-QT2-2021-04	352	0.0536	-	SFAM01.1	PPB	2/23/2021 13:33	SMPL
CCV020	CCV020	20656	5.0451	-	SFAM01.1	PPB	2/23/2021 13:35	SMPL
CCB020	CCB020	89	-0.0111	-	SFAM01.1	PPB	2/23/2021 13:38	SMPL