


LB 113227

7470A

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 2.3691e-004

C= 1.1925e-002

Rho= 0.9998973

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	% D
0.0	0.000	0.045	0.045	139	0.000	139					
0.2	0.200	0.196	-0.004	778	0.0 %	778					-2
2.5	2.500	2.522	0.022	10594	0.0 %	10594					1
5.0	5.000	4.903	-0.097	20646	0.0 %	20646					-2
7.5	7.500	7.471	-0.029	31483	0.0 %	31483					0
10.0	10.000	10.063	0.063	42428	0.0 %	42428					1

LB113227 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 23 Feb 2021 14:38:51

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	139	-	0.0000	7470A	PPB	23 Feb 2021 14:39:36	S	Std
0.2 - 1	50.2	778	-	0.2000	7470A	PPB	23 Feb 2021 14:41:53	S	Std
2.5 - 1	52.5	10594	-	2.5000	7470A	PPB	23 Feb 2021 14:44:44	S	Std
5.0 - 1	55	20646	-	5.0000	7470A	PPB	23 Feb 2021 14:47:01	S	Std
7.5 - 1	57.5	31483	-	7.5000	7470A	PPB	23 Feb 2021 14:49:18	S	Std
10.0 - 1	511	42428	-	10.0000	7470A	PPB	23 Feb 2021 14:51:34	S	Std
ICV10 - 1	ICV10	15440	3.6698	-	7470A	PPB	23 Feb 2021 14:54:27	U	SMPL
ICB10 - 1	ICB10	98	0.0351	-	7470A	PPB	23 Feb 2021 14:56:42	U	SMPL
CCV21 - 1	CCV21	20825	4.9456	-	7470A	PPB	23 Feb 2021 14:58:58	U	SMPL
CCB21 - 1	CCB21	18	0.0162	-	7470A	PPB	23 Feb 2021 15:01:13	U	SMPL
CRA - 1	CRA	759	0.1917	-	7470A	PPB	23 Feb 2021 15:03:07	U	SMPL
HighStd - 1	HighStd	40356	9.5726	-	7470A	PPB	23 Feb 2021 15:05:22	U	SMPL
ChkStd - 1	ChkStd	29691	7.0460	-	7470A	PPB	23 Feb 2021 15:07:37	U	SMPL
PB134763BL - 1	PBW	8	0.0138	-	7470A	PPB	23 Feb 2021 15:09:53	U	SMPL
PB134763BS - 1	LCSW	15424	3.6660	-	7470A	PPB	23 Feb 2021 15:14:46	U	SMPL
M1400-06 - 1	BIN-2	71	0.0287	-	7470A	PPB	23 Feb 2021 15:17:02	U	SMPL
M1400-06DUP - 1	BIN-2DUP	106	0.0370	-	7470A	PPB	23 Feb 2021 15:19:18	U	SMPL
M1400-06MS - 1	BIN-2MS	14679	3.4895	-	7470A	PPB	23 Feb 2021 15:23:22	U	SMPL
M1400-06MSD - 1	BIN-2MSD	14766	3.5101	-	7470A	PPB	23 Feb 2021 15:25:38	U	SMPL
M1452-04 - 1	TP-1	141	0.0453	-	7470A	PPB	23 Feb 2021 15:27:54	U	SMPL
CCV22 - 1	CCV22	21042	4.9970	-	7470A	PPB	23 Feb 2021 15:30:45	U	SMPL
CCB22 - 1	CCB22	126	0.0418	-	7470A	PPB	23 Feb 2021 15:32:59	U	SMPL
M1452-08 - 1	TP-2	166	0.0513	-	7470A	PPB	23 Feb 2021 15:35:15	U	SMPL
M1452-12 - 1	TP-3	165	0.0510	-	7470A	PPB	23 Feb 2021 15:37:30	U	SMPL
M1452-16 - 1	TP-4	153	0.0482	-	7470A	PPB	23 Feb 2021 15:39:46	U	SMPL
M1452-20 - 1	TP-5	129	0.0425	-	7470A	PPB	23 Feb 2021 15:42:02	U	SMPL
M1452-24 - 1	TP-6	481	0.1259	-	7470A	PPB	23 Feb 2021 15:44:18	U	SMPL
M1452-28 - 1	TP-7	131	0.0430	-	7470A	PPB	23 Feb 2021 15:46:35	U	SMPL
PB134763TB - 1	PB134749TB	136	0.0441	-	7470A	PPB	23 Feb 2021 15:48:52	U	SMPL
M1400-06LX5 - 1		131	0.0430	-	7470A	PPB	23 Feb 2021 15:51:09	U	SMPL
M1400-06A - 1		15675	3.7255	-	7470A	PPB	23 Feb 2021 15:53:26	U	SMPL
CCV23 - 1	CCV23	20847	4.9508	-	7470A	PPB	23 Feb 2021 15:55:42	U	SMPL
CCB23 - 1	CCB23	128	0.0422	-	7470A	PPB	23 Feb 2021 15:58:33	U	SMPL