



Analytical Summary Report

Analysis Method: 9040C

Analyst By : Alexander

Parameter: pH

Supervisor Review By : apatel

Run Number: LB113228

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2633
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.4	4.01	02/23/2021	14:50
2	CAL2	1	Water	NA	NA	20.4	7.01	02/23/2021	14:51
3	CAL3	1	Water	NA	NA	20.4	10.02	02/23/2021	14:52
4	ICV	1	Water	NA	NA	20.4	7.00	02/23/2021	14:55
5	CCV1	1	Water	NA	NA	20.3	2.01	02/23/2021	14:57
6	M1470-01	1	Water	NA	NA	18.6	6.94	02/23/2021	15:00
7	M1470-01DUP	1	Water	NA	NA	18.6	6.97	02/23/2021	15:02
8	M1470-02	1	Water	NA	NA	18.5	7.41	02/23/2021	15:05
9	M1470-03	1	Water	NA	NA	18.9	7.33	02/23/2021	15:07
10	M1470-06	1	Water	NA	NA	18.5	7.30	02/23/2021	15:10
11	CCV2	1	Water	NA	NA	20.3	12.02	02/23/2021	15:12

LB113228
WORKLIST(Hardcopy Internal Chain)

WorkList Name : pH-22321

WorkList ID : 146602

Department : Wet-Chemistry

Date : 02-23-2021 12:02:43

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/23/2021	Water	M1470-01 E	pH	Cool 4 deg C	TULL02	H31	CBH-594	02/23/2021	9040C
02/23/2021	Water	M1470-02 E	pH	Cool 4 deg C	TULL02	H31	CBH-591	02/23/2021	9040C
02/23/2021	Water	M1470-03 E	pH	Cool 4 deg C	TULL02	H31	CBH-139	02/23/2021	9040C
02/23/2021	Water	M1470-06 E	pH	Cool 4 deg C	TULL02	H31	CBH-140	02/23/2021	9040C

Date/Time 02/23/2021 14:40
Received by: AML(wet-chem)
Relinquished by: KS (wet-chem)

Date/Time 02/23/2021 15:40
Received by: KS (wet-chem)
Relinquished by: AML(wet-chem)