



Analytical Summary Report

Analysis Method: 9040C

Analyst By : Alexander

Parameter: pH

Supervisor Review By : apatel

Run Number: LB113240

Slope : 99.0

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2633
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	02/24/2021	11:20
2	CAL2	1	Water	NA	NA	20.3	7.01	02/24/2021	11:21
3	CAL3	1	Water	NA	NA	20.3	10.03	02/24/2021	11:23
4	ICV	1	Water	NA	NA	20.3	7.02	02/24/2021	11:26
5	CCV1	1	Water	NA	NA	20.4	2.01	02/24/2021	11:28
6	M1481-02	1	Water	NA	NA	18.5	6.45	02/24/2021	11:30
7	M1481-02DUP	1	Water	NA	NA	18.5	6.42	02/24/2021	11:32
8	M1481-03	1	Water	NA	NA	18.6	8.52	02/24/2021	11:35
9	M1481-04	1	Water	NA	NA	18.9	7.08	02/24/2021	11:38
10	CCV2	1	Water	NA	NA	20.4	12.01	02/24/2021	11:40

LB113240 WORKLIST(Hardcopy Internal Chain)

WorkList Name : pH-22421 WorkList ID : 146616 Department : Wet-Chemistry Date : 02-24-2021 10:09:15

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/23/2021	Water	M1481-02	pH	Cool 4 deg C	TETR06	H41	TT-025-IDWGW-20210223	02/23/2021	9040C
02/23/2021	Water	M1481-03	pH	Cool 4 deg C	TETR06	H41	TT-026-IDWGW-20210223	02/23/2021	9040C
02/23/2021	Water	M1481-04	pH	Cool 4 deg C	TETR06	H41	TT-027-IDWGW-20210223	02/23/2021	9040C

Date/Time 02/24/2021 11:10
 Received by: AM (wet-chem)
 Relinquished by: KS (wet-chem)

Date/Time 02/24/2021 11:58
 Received by: KS (wet-chem)
 Relinquished by: AM (wet-chem)