


LB113242

7471B

INSTRUMENT ID: CY1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.004	0.004	167	0.000	167					1
0.2	0.200	0.184	-0.016	1190	0.0 %	1190					-8
2.5	2.500	2.443	-0.057	14023	0.0 %	14023					-2
5.0	5.000	5.067	0.067	28931	0.0 %	28931					1
7.5	7.500	7.595	0.095	43287	0.0 %	43287					1
10.0	10.000	9.907	-0.093	56424	0.0 %	56424					1

LB113242 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 24 Feb 2021 12:25:37

Sample ID	Extended ID	μ Abs.	Conc.	Std. Conc	Method	Units	Date	Type	Type
0.0 - 1	50	167	-	0.0000	7471B	PPB	24 Feb 2021 12:29:20	S	Std
0.2 - 1	50.2	1190	-	0.2000	7471B	PPB	24 Feb 2021 12:31:38	S	Std
2.5 - 1	52.5	14023	-	2.5000	7471B	PPB	24 Feb 2021 12:33:55	S	Std
5.0 - 1	55	28931	-	5.0000	7471B	PPB	24 Feb 2021 12:36:12	S	Std
7.5 - 1	57.5	43287	-	7.5000	7471B	PPB	24 Feb 2021 12:40:59	S	Std
10.0 - 1	518	56424	-	10.0000	7471B	PPB	24 Feb 2021 12:43:15	S	Std
ICV12 - 1	ICV12	22229	3.8876	-	7471B	PPB	24 Feb 2021 12:46:39	U	SMPL
ICB12 - 1	ICB12	10	-0.0239	-	7471B	PPB	24 Feb 2021 12:48:54	U	SMPL
CCV27 - 1	CCV27	29081	5.0938	-	7471B	PPB	24 Feb 2021 12:51:09	U	SMPL
CCB27 - 1	CCB27	23	-0.0216	-	7471B	PPB	24 Feb 2021 12:53:25	U	SMPL
CRA - 1	CRA	1172	0.1807	-	7471B	PPB	24 Feb 2021 12:55:13	U	SMPL
HighStd - 1	HighStd	56886	9.9886	-	7471B	PPB	24 Feb 2021 12:59:16	U	SMPL
ChkStd - 1	ChkStd	41354	7.2544	-	7471B	PPB	24 Feb 2021 13:01:31	U	SMPL
PB134785BL - 1	PBS	-61	-0.0364	-	7471B	PPB	24 Feb 2021 13:03:48	U	SMPL
PB134785BS - 1	LCSS	21257	3.7165	-	7471B	PPB	24 Feb 2021 13:06:05	U	SMPL
M1460-01 - 1	HD-01-022221	788	0.1131	-	7471B	PPB	24 Feb 2021 13:08:22	U	SMPL
M1462-01 - 1	CL-01-022221	20027	3.4999	-	7471B	PPB	24 Feb 2021 13:10:38	U	SMPL
M1462-01DUP - 1	CL-01-022221DUP	22679	3.9668	-	7471B	PPB	24 Feb 2021 13:12:55	U	SMPL
M1462-01MS - 1	CL-01-022221MS	44806	7.8621	-	7471B	PPB	24 Feb 2021 13:15:13	U	SMPL
M1462-01MSD - 1	CL-01-022221MSD	43335	7.6031	-	7471B	PPB	24 Feb 2021 13:17:28	U	SMPL
CCV28 - 1	CCV28	29082	5.0940	-	7471B	PPB	24 Feb 2021 13:21:45	U	SMPL
CCB28 - 1	CCB28	-19	-0.0290	-	7471B	PPB	24 Feb 2021 13:24:00	U	SMPL
M1462-01LX5 - 1		3949	0.6696	-	7471B	PPB	24 Feb 2021 13:26:16	U	SMPL
M1462-01A - 1		43873	7.6978	-	7471B	PPB	24 Feb 2021 13:28:31	U	SMPL
CCV29 - 1	CCV29	29268	5.1267	-	7471B	PPB	24 Feb 2021 13:30:47	U	SMPL
CCB29 - 1	CCB29	75	-0.0124	-	7471B	PPB	24 Feb 2021 13:33:32	U	SMPL