



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB113246

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID:

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.00	80.6	80.6	02/24/2021	14:10
2	M1470-01		1	9	100.00	>212.0	>212.0	02/24/2021	14:40
3	M1470-01DUP		1	9	100.00	>212.0	>212.0	02/24/2021	15:05
4	M1470-02		1	11	100.00	>212.0	>212.0	02/24/2021	15:35
5	M1470-03		1	10	100.00	>212.0	>212.0	02/24/2021	16:05
6	M1470-06		1	11	100.00	>212.0	>212.0	02/24/2021	16:30

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP22421 WorkList ID : 146628 Department : Wet-Chemistry Date : 02-24-2021 13:11:13

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/02/2021	Water	M1470-01	Flash Point	Cool 4 deg C	TULL02	H31	CBH-594	02/23/2021	1010A
03/02/2021	Water	M1470-02	Flash Point	Cool 4 deg C	TULL02	H31	CBH-591	02/23/2021	1010A
03/02/2021	Water	M1470-03	Flash Point	Cool 4 deg C	TULL02	H31	CBH-139	02/23/2021	1010A
03/02/2021	Water	M1470-06	Flash Point	Cool 4 deg C	TULL02	H31	CBH-140	02/23/2021	1010A

Date/Time 02/24/21 13:45
 Received by: KS (Welcher)
 Relinquished by: KS (Welcher)

Date/Time 02/24/21 16:35
 Received by: KS (Welcher)
 Relinquished by: KS (Welcher)