


LB 113259

7470A



Linear

A= 0.0000e+000

B= 2.3822e-004

C= -2.0728e-002

Rho= 0.9999101

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/0 D
0.0	0.000	-0.006	-0.006	61	0.000	61					-10
0.2	0.200	0.181	-0.019	846	0.0 %	846					3
2.5	2.500	2.578	0.078	10909	0.0 %	10909					-2
5.0	5.000	4.918	-0.082	20730	0.0 %	20730					0
7.5	7.500	7.537	0.037	31725	0.0 %	31725					0
10.0	10.000	9.993	-0.007	42036	0.0 %	42036					

LB113259 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 25 Feb 2021 11:20:37

Sample ID	Extended ID	μ Abs	Conc.	Std CondMethod	Units	Date	Type	Type
0.0 - 1	SO	61	-	0.00007470A	PPB	25 Feb 2021 11:26:28	S	Std
0.2 - 1	SO.2	846	-	0.20007470A	PPB	25 Feb 2021 11:28:45	S	Std
2.5 - 1	SO.5	10909	-	2.50007470A	PPB	25 Feb 2021 11:31:02	S	Std
5.0 - 1	SO	20730	-	5.00007470A	PPB	25 Feb 2021 11:33:19	S	Std
7.5 - 1	SO.5	31725	-	7.50007470A	PPB	25 Feb 2021 11:37:21	S	Std
10.0 - 1	SO	42036	-	10.00007470A	PPB	25 Feb 2021 11:39:37	S	Std
ICV13 - 1	ICV13	15947	3.7782	-7470A	PPB	25 Feb 2021 11:42:31	U	SMPL
ICB13 - 1	ICB13	-27	-0.0272	-7470A	PPB	25 Feb 2021 11:44:46	U	SMPL
CCV30 - 1	CCV30	20739	4.9197	-7470A	PPB	25 Feb 2021 11:47:01	U	SMPL
CCB30 - 1	CCB30	-22	-0.0260	-7470A	PPB	25 Feb 2021 11:49:16	U	SMPL
CRA - 1	CRA	832	0.1775	-7470A	PPB	25 Feb 2021 11:51:31	U	SMPL
HighStd - 1	HighStd	42060	9.9988	-7470A	PPB	25 Feb 2021 11:53:46	U	SMPL
ChkStd - 1	ChkStd	30132	7.1573	-7470A	PPB	25 Feb 2021 11:56:02	U	SMPL
PB134815BL - 1	PBW	-70	-0.0374	-7470A	PPB	25 Feb 2021 11:58:18	U	SMPL
PB134815BS - 1	LCSW	17637	4.1807	-7470A	PPB	25 Feb 2021 12:00:33	U	SMPL
M1481-02 - 1	TT-025-IDWGW-20210223	12	-0.0179	-7470A	PPB	25 Feb 2021 12:02:50	U	SMPL
M1481-02DUP - 1	TT-025-IDWGW-20210223DUP	60	-0.0064	-7470A	PPB	25 Feb 2021 12:05:06	U	SMPL
M1481-02MS - 1	TT-025-IDWGW-20210223MS	18134	4.2991	-7470A	PPB	25 Feb 2021 12:07:23	U	SMPL
M1481-02MSD - 1	TT-025-IDWGW-20210223MSD	16363	3.8773	-7470A	PPB	25 Feb 2021 12:09:41	U	SMPL
M1481-03 - 1	TT-026-IDWGW-20210223	54	-0.0079	-7470A	PPB	25 Feb 2021 12:11:56	U	SMPL
CCV31 - 1	CCV31	20600	4.8866	-7470A	PPB	25 Feb 2021 12:14:11	U	SMPL
CCB31 - 1	CCB31	-37	-0.0295	-7470A	PPB	25 Feb 2021 12:16:27	U	SMPL
M1481-04 - 1	TT-027-IDWGW-20210223	137	0.0119	-7470A	PPB	25 Feb 2021 12:18:42	U	SMPL
M1481-02LX5 - 1		34	-0.0126	-7470A	PPB	25 Feb 2021 12:20:57	U	SMPL
M1481-02A - 1		17167	4.0688	-7470A	PPB	25 Feb 2021 12:23:13	U	SMPL
CCV32 - 1	CCV32	20597	4.8859	-7470A	PPB	25 Feb 2021 12:25:29	U	SMPL
CCB32 - 1	CCB32	-38	-0.0298	-7470A	PPB	25 Feb 2021 12:27:45	U	SMPL