


LB113263

7471B

INSTRUMENT ID: CV1

μ Abs.:

54958



Linear

A= 0.0000e+000

B= 1.8220e-004

C= 1.0922e-002

Rho= 0.9999692

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%
0.0	0.000	0.019	0.019	47	0.000	47					
0.2	0.200	0.226	0.026	1181	0.0 %	1181					
2.5	2.500	2.470	-0.030	13499	0.0 %	13499					
5.0	5.000	4.951	-0.049	27112	0.0 %	27112					
7.5	7.500	7.509	0.009	41151	0.0 %	41151					
10.0	10.000	10.024	0.024	54958	0.0 %	54958					

0/0
13
11
100

LB113263 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 25 Feb 2021 13:48:15

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	47	-	0.0000	7471B	PPB	25 Feb 2021 14:10:08	S	Std
0.2 - 1	50.2	1181	-	0.2000	7471B	PPB	25 Feb 2021 14:12:25	S	Std
2.5 - 1	52.5	13499	-	2.5000	7471B	PPB	25 Feb 2021 14:14:43	S	Std
5.0 - 1	55	27112	-	5.0000	7471B	PPB	25 Feb 2021 14:16:59	S	Std
7.5 - 1	57.5	41151	-	7.5000	7471B	PPB	25 Feb 2021 14:19:16	S	Std
10.0 - 1	511	54958	-	10.0000	7471B	PPB	25 Feb 2021 14:24:07	S	Std
ICV14 - 1	ICV14	21845	3.9911	-	7471B	PPB	25 Feb 2021 14:26:52	U	SMPL
ICB14 - 1	ICB14	-65	-0.0009	-	7471B	PPB	25 Feb 2021 14:29:07	U	SMPL
CCV33 - 1	CCV33	26760	4.8867	-	7471B	PPB	25 Feb 2021 14:31:22	U	SMPL
CCB33 - 1	CCB33	-48	0.0022	-	7471B	PPB	25 Feb 2021 14:33:37	U	SMPL
CRA - 1	CRA	1152	0.2208	-	7471B	PPB	25 Feb 2021 14:35:53	U	SMPL
HighStd - 1	HighStd	54759	9.9881	-	7471B	PPB	25 Feb 2021 14:38:08	U	SMPL
ChkStd - 1	ChkStd	39648	7.2349	-	7471B	PPB	25 Feb 2021 14:40:24	U	SMPL
PB134822BL - 1	PBS	-55	0.0009	-	7471B	PPB	25 Feb 2021 14:42:41	U	SMPL
PB134822BS - 1	LCSS	21340	3.8991	-	7471B	PPB	25 Feb 2021 14:44:57	U	SMPL
M1469-01 - 1	TR-08-022321	6600	1.2135	-	7471B	PPB	25 Feb 2021 14:47:14	U	SMPL
M1469-01DUP - 1	TR-08-022321DUP	5968	1.0983	-	7471B	PPB	25 Feb 2021 14:49:30	U	SMPL
M1469-01MS - 1	TR-08-022321MS	27418	5.0065	-	7471B	PPB	25 Feb 2021 14:51:47	U	SMPL
M1469-01MSD - 1	TR-08-022321MSD	28774	5.2536	-	7471B	PPB	25 Feb 2021 14:54:04	U	SMPL
M1472-03 - 1	COMPOSITE-01(TP2-TP1)	5173	0.9535	-	7471B	PPB	25 Feb 2021 14:56:20	U	SMPL
CCV34 - 1	CCV34	26889	4.9102	-	7471B	PPB	25 Feb 2021 14:58:35	U	SMPL
CCB34 - 1	CCB34	-41	0.0035	-	7471B	PPB	25 Feb 2021 15:00:51	U	SMPL
M1477-01 - 1	72-11995	311	0.0676	-	7471B	PPB	25 Feb 2021 15:03:06	U	SMPL
M1469-01LX5 - 1		1274	0.2430	-	7471B	PPB	25 Feb 2021 15:05:22	U	SMPL
M1469-01A - 1		28135	5.1372	-	7471B	PPB	25 Feb 2021 15:07:37	U	SMPL
CCV35 - 1	CCV35	27080	4.9450	-	7471B	PPB	25 Feb 2021 15:09:53	U	SMPL
CCB35 - 1	CCB35	-52	0.0014	-	7471B	PPB	25 Feb 2021 15:12:09	U	SMPL