



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Alexander

Parameter: pH

Supervisor Review By : apatel

Run Number: LB113267

Slope : 99.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2633
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	02/25/2021	15:58
2	CAL2	1	Water	NA	NA	20.6	7.01	02/25/2021	15:59
3	CAL3	1	Water	NA	NA	20.6	10.02	02/25/2021	16:01
4	ICV	1	Water	NA	NA	20.5	7.02	02/25/2021	16:04
5	CCV1	1	Water	NA	NA	20.6	2.01	02/25/2021	16:06
6	M1503-01	1	Solid	20.08	20	20.3	6.75	02/25/2021	16:08
7	M1503-01DUP	1	Solid	20.08	20	20.4	6.74	02/25/2021	16:10
8	CCV2	1	Water	NA	NA	20.7	12.01	02/25/2021	16:13

LB13267

WORKLIST(Hardcopy Internal Chain)

WorkList Name : pH-soil-022521 WorkList ID : 146671 Department : Wet-Chemistry Date : 02-25-2021 15:31:57

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/25/2021	Solid	M1503-01	pH	Cool 4 deg C	VERI01	G61	SEDGWICK-AVE	02/25/2021	9045D

Date/Time 02/25/2021 15:50
Received by: AMCwet-chem
Relinquished by: KS (Welch)

Date/Time 02/25/2021 17:06
Received by: KS (Welch)
Relinquished by: AMCwet-chem