



PERCENT SOLID

Supervisor: apatel
Analyst: Alexander
Date: 2/26/2021

OVENTEMP IN Celsius(°C): 107
Time IN: 16:50
In Date: 02/25/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 106
Time OUT: 09:10
Out Date: 02/26/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB113269

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
M1490-01	02-25-A	1	1.00	2.00	2.00	100.0	plastic sample
M1490-02	02-25-B	2	1.00	2.00	2.00	100.0	plastic sample
M1490-03	02-25-C	3	1.00	2.00	2.00	100.0	plastic sample
M1490-04	02-25-SOIL	4	1.00	2.00	2.00	100.0	STONE SAMPLE,100% SOLIDS
M1491-01	SA-01-022521	5	1.15	9.83	8.8	88.1	
M1491-02	SA-01-022521-EPH-2	6	1.14	9.94	8.85	87.6	
M1491-02D	SA-01-022521-EPH-2DUP	7	1.15	9.98	8.9	87.8	
M1497-01	WASTE	8	1.14	9.54	8.63	89.2	
M1499-01	N-1	9	1.17	9.9	9.32	93.4	
M1499-02	N-1A	10	1.17	9.95	9.47	94.5	
M1500-01	T-1	11	1.16	9.72	8.93	90.8	
M1500-02	T-2	12	1.15	9.6	8.56	87.7	
M1500-03	T-3	13	1.17	9.65	8.7	88.8	
M1500-04	T-4	14	1.14	9.93	8.8	87.1	
M1500-05	T-5	15	1.14	9.74	8.49	85.5	
M1503-01	SEDGWICK-AVE	16	1.17	9.52	8.65	89.6	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$