

TOTAL SOLIDS - SM2540B

SUPERVISOR: apatel

ANALYST: JIGNESH

Date: 02/26/2021

OVEN IN TEMP1: 103 °C

OVEN OUT TEMP1: 104 °C

DATE & TIME IN1: 02/25/2021 16:00 DATE & TIME OUT1: 02/26/2021 08:10

OVEN IN TEMP2: 104 °C

OVEN OUT TEMP2: 104 °C

DATE & TIME IN2: 02/26/2021 08:45 DATE & TIME OUT2: 02/26/2021 10:15

Run Number: LB113276

BalanceID: WC SC-3

OvenID: WC OVEN-1

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish+Sample weight after Drying @103-@105°C (g)	2nd Dish+Sample weight after Drying @103-@105°C (g)	Final Dish+Sample weight after Drying @103-@105°C (g)	Weight (g)	Result (mg/L)
1	LB113276BL	LB113276BL	14.8971	14.8971	100	14.8971	14.8971	14.8971	0.0000	0
2	M1475-05	WA-CON	156.5636	156.5636	100	156.5844	156.5844	156.5844	0.0208	208
3	M1475-05DUP	WA-CONDUP	151.2240	151.2240	100	151.2438	151.2438	151.2438	0.0198	198
4	M1476-05	ML-CON	157.5164	157.5164	100	157.5270	157.5270	157.5270	0.0106	106

A = Sample Volume (ml)

B = Final Empty Dish Weight (g)

C = Final Dish+Sample weight after Drying @103-@105°C (g)

$$\text{Result mg/L} = ((C - B) / A) * 1000 * 1000$$

LB113276 WORKLIST(Hardcopy Internal Chain)

WorkList Name : ASTM-TS-22521

WorkList ID : 146654

Department : Wet-Chemistry

Date : 02-25-2021 10:11:59

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/03/2021	Solid	M1475-05	ASTM TS	Cool 4 deg C	PSEG03	I21	WA-CON	02/24/2021	SM2540 B
03/03/2021	Solid	M1476-05	ASTM TS	Cool 4 deg C	PSEG03	I21	ML-CON	02/24/2021	SM2540 B

Date/Time 02-25-2021 14:00
Received by: AMCWetchem
Relinquished by: KS (Wetchem)

Date/Time 02-25-2021 16:40
Received by: KS (Wetchem)
Relinquished by: AMCWetchem