



Analytical Summary Report

Analysis Method: 9040C

Analyst By : Alexander

Parameter: pH

Supervisor Review By : apatel

Run Number: LB113277

Slope : 99.0

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2633
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	02/26/2021	09:05
2	CAL2	1	Water	NA	NA	20.8	7.01	02/26/2021	09:07
3	CAL3	1	Water	NA	NA	20.7	10.02	02/26/2021	09:08
4	ICV	1	Water	NA	NA	20.6	7.00	02/26/2021	09:12
5	CCV1	1	Water	NA	NA	20.6	2.01	02/26/2021	09:14
6	M1488-04	1	Water	NA	NA	20.2	6.27	02/26/2021	09:16
7	M1488-04DUP	1	Water	NA	NA	20.2	6.25	02/26/2021	09:18
8	M1488-05	1	Water	NA	NA	20.3	10.14	02/26/2021	09:21
9	M1488-06	1	Water	NA	NA	20.2	6.38	02/26/2021	09:23
10	M1488-07	1	Water	NA	NA	20.1	5.43	02/26/2021	09:26
11	M1489-04	1	Water	NA	NA	20.4	6.07	02/26/2021	09:29
12	M1489-05	1	Water	NA	NA	20.0	6.02	02/26/2021	09:31
13	M1489-07	1	Water	NA	NA	20.1	5.38	02/26/2021	09:33
14	CCV2	1	Water	NA	NA	20.6	12.01	02/26/2021	09:35

LB113277 WORKLIST(Hardcopy Internal Chain)

WorkList Name : pH-water-022521-2 WorkList ID : 146673 Department : Wet-Chemistry Date : 02-25-2021 15:32:09

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/25/2021	Water	M1488-04	pH	Cool 4 deg C	PSEG03	I32	41709	02/25/2021	9040C
02/25/2021	Water	M1488-05	pH	Cool 4 deg C	PSEG03	I32	41698	02/25/2021	9040C
02/25/2021	Water	M1488-06	pH	Cool 4 deg C	PSEG03	I32	41697	02/25/2021	9040C
02/25/2021	Water	M1488-07	pH	Cool 4 deg C	PSEG03	I32	30779	02/25/2021	9040C
02/25/2021	Water	M1489-04	pH	Cool 4 deg C	PSEG03	I33	COMP-41071-72	02/25/2021	9040C
02/25/2021	Water	M1489-05	pH	Cool 4 deg C	PSEG03	I33	COMP-41706-07	02/25/2021	9040C
02/25/2021	Water	M1489-07	pH	Cool 4 deg C	PSEG03	I33	41108	02/25/2021	9040C

Date/Time 02/26/2021 9:00
 Received by: APM (wet-chem)
 Relinquished by: KS (wet-chem)

Date/Time 02/26/2021 9:50
 Received by: KS (wet-chem)
 Relinquished by: APM (wet-chem)