


LB113281

245.1

INSTRUMENT ID: CV1

 μ Abs.:

40389



Linear

A= 0.0000e+000

B= 2.4841e-004

C= -9.5594e-003

Rho= 0.9999616

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.001	-0.001	36	0.000	36					-2
0.05	0.050	0.049	-0.001	234	0.0 %	234	0				-7
0.2	0.200	0.187	-0.013	790	0.0 %	790					1
2.5	2.500	2.513	0.013	10155	0.0 %	10155					1
5.0	5.000	5.045	0.045	20349	0.0 %	20349					-1
7.5	7.500	7.433	-0.067	29962	0.0 %	29962					0
10.0	10.000	10.024	0.024	40389	0.0 %	40389					

LB113281 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 26 Feb 2021 11:52:18

Sample ID	Extended ID	μ Abs.	Conc.	Strnd Conc	Method	Units	Date	Type	Type
0.0 - 1	SO	36	-	0.0000	245.1	PPB	26 Feb 2021 11:54:33	S	Std
0.05 - 1	SO.05	234	-	0.0500	245.1	PPB	26 Feb 2021 11:56:50	S	Std
0.2 - 1	SO.2	790	-	0.2000	245.1	PPB	26 Feb 2021 11:59:07	S	Std
2.5 - 1	SO.5	10155	-	2.5000	245.1	PPB	26 Feb 2021 12:01:24	S	Std
5.0 - 1	SO.5	20349	-	5.0000	245.1	PPB	26 Feb 2021 12:03:41	S	Std
7.5 - 1	SO.5	29962	-	7.5000	245.1	PPB	26 Feb 2021 12:05:57	S	Std
10.0 - 1	SO.5	40389	-	10.0000	245.1	PPB	26 Feb 2021 12:10:48	S	Std
ICV16 - 1	ICV16	16354	4.0530	-245.1		PPB	26 Feb 2021 12:16:10	U	SMPL
ICB16 - 1	ICB16	-24	-0.0155	-245.1		PPB	26 Feb 2021 12:18:25	U	SMPL
CCV39 - 1	CCV39	20200	5.0084	-245.1		PPB	26 Feb 2021 12:20:40	U	SMPL
CCB39 - 1	CCB39	-48	-0.0215	-245.1		PPB	26 Feb 2021 12:22:55	U	SMPL
CRA - 1	CRA	217	0.0443	-245.1		PPB	26 Feb 2021 12:27:45	U	SMPL
HighStd - 1	HighStd	40661	10.0911	-245.1		PPB	26 Feb 2021 12:30:00	U	SMPL
ChkStd - 1	ChkStd	27888	6.9182	-245.1		PPB	26 Feb 2021 12:32:15	U	SMPL
PB134827BL - 1	PBW	-80	-0.0294	-245.1		PPB	26 Feb 2021 12:34:32	U	SMPL
PB134827BS - 1	LCSW	15086	3.7380	-245.1		PPB	26 Feb 2021 12:36:48	U	SMPL
M1470-01 - 1	CBH-594	-19	-0.0143	-245.1		PPB	26 Feb 2021 12:39:04	U	SMPL
M1470-01DUP - 1	CBH-594DUP	23	-0.0038	-245.1		PPB	26 Feb 2021 12:41:21	U	SMPL
M1470-01MS - 1	CBH-594MS	14966	3.7082	-245.1		PPB	26 Feb 2021 12:43:38	U	SMPL
M1470-01MSD - 1	CBH-594MSD	14497	3.5917	-245.1		PPB	26 Feb 2021 12:45:55	U	SMPL
M1470-02 - 1	CBH-591	-20	-0.0145	-245.1		PPB	26 Feb 2021 12:48:11	U	SMPL
CCV40 - 1	CCV40	19923	4.9396	-245.1		PPB	26 Feb 2021 12:50:26	U	SMPL
CCB40 - 1	CCB40	-33	-0.0178	-245.1		PPB	26 Feb 2021 12:52:41	U	SMPL
M1470-03 - 1	CBH-139	42	0.0009	-245.1		PPB	26 Feb 2021 12:54:57	U	SMPL
M1470-06 - 1	CBH-140	64	0.0063	-245.1		PPB	26 Feb 2021 12:57:13	U	SMPL
M1470-07 - 1	CBH-594	19	-0.0048	-245.1		PPB	26 Feb 2021 12:59:28	U	SMPL
M1470-08 - 1	CBH-591	19	-0.0048	-245.1		PPB	26 Feb 2021 13:01:44	U	SMPL
M1470-09 - 1	CBH-139	30	-0.0021	-245.1		PPB	26 Feb 2021 13:04:00	U	SMPL
M1470-10 - 1	CBH-140	42	0.0009	-245.1		PPB	26 Feb 2021 13:06:17	U	SMPL
M1470-01LX5 - 1		14	-0.0061	-245.1		PPB	26 Feb 2021 13:08:34	U	SMPL
M1470-01A - 1		18021	4.4671	-245.1		PPB	26 Feb 2021 13:10:51	U	SMPL
CCV41 - 1	CCV41	20188	5.0054	-245.1		PPB	26 Feb 2021 13:13:09	U	SMPL
CCB41 - 1	CCB41	-40	-0.0195	-245.1		PPB	26 Feb 2021 13:15:25	U	SMPL